



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB102375

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/26/2019	10:35
2	K2578-02		1	8	100.00	>212.0	>212.0	04/26/2019	11:05
3	K2578-02DUP		1	8	100.00	>212.0	>212.0	04/26/2019	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102375

WorkList Name : FP42619

WorkList ID : 125354

Date : 04-26-2019 09:03:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2019	Water	K2578-02	Flash Point	Cool 4 deg C	PSEG01	K11	WATER-COMP	04/25/2019	1010A

Date/Time 04/26/19 10:05
 Received by: KS
 Relinquished by: CF

Date/Time 04/26/19 11:45
 Received by: CF
 Relinquished by: KS