

LB102380


7470A

INSTRUMENT ID : CUI

Linear

 μ Abs. :

4542E



A= 0.0000e+000

B= 2.1889e-004

C= -2.6676e-002

Rho= 0.9990931

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	96	0.000	96				
0.2	0.200	0.162	-0.038	864	0.0 %	864				
2.5	2.500	2.381	-0.119	10998	0.0 %	10998				
5.0	5.000	5.341	0.341	24523	0.0 %	24523				
7.5	7.500	7.405	-0.095	33950	0.0 %	33950				
10.0	10.000	9.917	-0.083	45426	0.0 %	45426				

LB102380 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Apr 2019 14:23:33

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	96	-	0.0000	7470A	PPB	26 Apr 2019 14:25:44	S	Std
0.2 - 1	50.2	864	-	0.2000	7470A	PPB	26 Apr 2019 14:27:52	S	Std
2.5 - 1	52.5	10998	-	2.5000	7470A	PPB	26 Apr 2019 14:30:00	S	Std
5.0 - 1	55.0	24523	-	5.0000	7470A	PPB	26 Apr 2019 14:32:08	S	Std
7.5 - 1	57.5	33950	-	7.5000	7470A	PPB	26 Apr 2019 14:34:15	S	Std
10.0 - 1	510.0	45426	-	10.0000	7470A	PPB	26 Apr 2019 14:36:22	S	Std
ICV22 - 1	ICV22	17999	3.9131	-	7470A	PPB	26 Apr 2019 14:40:26	U	SMPL
ICB22 - 1	ICB22	0	-0.0267	-	7470A	PPB	26 Apr 2019 14:42:32	U	SMPL
CCV03 - 1	CCV03	22588	4.9176	-	7470A	PPB	26 Apr 2019 14:44:38	U	SMPL
CCB03 - 1	CCB03	-17	-0.0304	-	7470A	PPB	26 Apr 2019 14:46:45	U	SMPL
CRA - 1	CRA	808	0.1502	-	7470A	PPB	26 Apr 2019 14:48:51	U	SMPL
HighStd - 1	HighStd	44801	9.7799	-	7470A	PPB	26 Apr 2019 14:50:56	U	SMPL
ChkStd - 1	ChkStd	31681	6.9080	-	7470A	PPB	26 Apr 2019 14:53:03	U	SMPL
PB119246BL - 1	PBW	-79	-0.0440	-	7470A	PPB	26 Apr 2019 14:55:10	U	SMPL
PB119246BS - 1	LCSW	18546	4.0329	-	7470A	PPB	26 Apr 2019 14:57:17	U	SMPL
K2203-18 - 1	SU-03-042419-A	-78	-0.0437	-	7470A	PPB	26 Apr 2019 14:59:24	U	SMPL
K2203-18DUP - 1	SU-03-042419-ADUP	-55	-0.0387	-	7470A	PPB	26 Apr 2019 15:01:31	U	SMPL
K2203-18MS - 1	SU-03-042419-AMS	19112	4.1568	-	7470A	PPB	26 Apr 2019 15:03:39	U	SMPL
K2203-18MSD - 1	SU-03-042419-AMSD	18961	4.1237	-	7470A	PPB	26 Apr 2019 15:05:46	U	SMPL
K2203-20 - 1	SU-03-042419-B	-31	-0.0335	-	7470A	PPB	26 Apr 2019 15:07:53	U	SMPL
CCV04 - 1	CCV04	22375	4.8710	-	7470A	PPB	26 Apr 2019 15:09:59	U	SMPL
CCB04 - 1	CCB04	-6	-0.0280	-	7470A	PPB	26 Apr 2019 15:12:05	U	SMPL
K2544-02 - 1	135-20	9	-0.0247	-	7470A	PPB	26 Apr 2019 15:14:11	U	SMPL
K2553-03 - 1	RBR200034	5	-0.0256	-	7470A	PPB	26 Apr 2019 15:16:16	U	SMPL
K2553-06 - 1	RBR200051	14	-0.0236	-	7470A	PPB	26 Apr 2019 15:18:23	U	SMPL
K2554-03 - 1	RT3860	30	-0.0201	-	7470A	PPB	26 Apr 2019 15:20:29	U	SMPL
PB119247BL - 1	PBW	43	-0.0173	-	7470A	PPB	26 Apr 2019 15:22:36	U	SMPL
PB119247BS - 1	LCSW	18140	3.9440	-	7470A	PPB	26 Apr 2019 15:24:43	U	SMPL
K2538-06 - 1	NB-20523	72	-0.0109	-	7470A	PPB	26 Apr 2019 15:26:51	U	SMPL
K2541-01 - 1	DRUM-COMP	100	-0.0048	-	7470A	PPB	26 Apr 2019 15:28:58	U	SMPL
K2545-02 - 1	FMI-133	27	-0.0208	-	7470A	PPB	26 Apr 2019 15:31:06	U	SMPL
K2545-02DUP - 1	FMI-133DUP	25	-0.0212	-	7470A	PPB	26 Apr 2019 15:33:13	U	SMPL
CCV05 - 1	CCV05	21860	4.7583	-	7470A	PPB	26 Apr 2019 15:35:20	U	SMPL
CCB05 - 1	CCB05	16	-0.0232	-	7470A	PPB	26 Apr 2019 15:37:30	U	SMPL
K2545-02MS - 1	FMI-133MS	19065	4.1465	-	7470A	PPB	26 Apr 2019 15:39:39	U	SMPL
K2545-02MSD - 1	FMI-133MSD	18939	4.1189	-	7470A	PPB	26 Apr 2019 15:41:50	U	SMPL
PB119248BL - 1	PBW	-36	-0.0346	-	7470A	PPB	26 Apr 2019 15:43:58	U	SMPL
PB119248BS - 1	LCSW	18224	3.9624	-	7470A	PPB	26 Apr 2019 15:46:04	U	SMPL
K2198-08 - 1	OK-01-042519	-31	-0.0335	-	7470A	PPB	26 Apr 2019 15:48:13	U	SMPL
K2198-08DUP - 1	OK-01-042519DUP	5	-0.0256	-	7470A	PPB	26 Apr 2019 15:50:23	U	SMPL
K2198-08MS - 1	OK-01-042519MS	17567	3.8186	-	7470A	PPB	26 Apr 2019 15:52:34	U	SMPL
K2198-08MSD - 1	OK-01-042519MSD	17753	3.8593	-	7470A	PPB	26 Apr 2019 15:54:48	U	SMPL
K2570-01 - 1	1-CB-NORTH	54	-0.0149	-	7470A	PPB	26 Apr 2019 15:56:57	U	SMPL
K2570-02 - 1	2-CB-SOUTH	-9	-0.0286	-	7470A	PPB	26 Apr 2019 15:59:03	U	SMPL
CCV06 - 1	CCV06	22749	4.9529	-	7470A	PPB	26 Apr 2019 16:01:11	U	SMPL
CCB06 - 1	CCB06	-48	-0.0372	-	7470A	PPB	26 Apr 2019 16:03:17	U	SMPL
K2575-02 - 1	PAINT-MATERIAL-DRUM	-19	-0.0308	-	7470A	PPB	26 Apr 2019 16:05:25	U	SMPL
PB119249BL - 1	PBW	549	0.0935	-	7470A	PPB	26 Apr 2019 16:07:32	U	SMPL
PB119249BS - 1	LCSW	18309	3.9810	-	7470A	PPB	26 Apr 2019 16:09:39	U	SMPL
K2545-01 - 1	FMI-133	3	-0.0260	-	7470A	PPB	26 Apr 2019 16:11:46	U	SMPL
K2545-01DUP - 1	FMI-133DUP	54	-0.0149	-	7470A	PPB	26 Apr 2019 16:13:53	U	SMPL
K2545-01MS - 1	FMI-133MS	18774	4.0828	-	7470A	PPB	26 Apr 2019 16:16:00	U	SMPL
K2545-01MSD - 1	FMI-133MSD	18583	4.0410	-	7470A	PPB	26 Apr 2019 16:18:07	U	SMPL
K2578-01 - 1	WATER-COMP	50204	10.9625	-	7470A	PPB	26 Apr 2019 16:20:15	U	SMPL
PB119246TB - 1	PB119196TB	-117	-0.0523	-	7470A	PPB	26 Apr 2019 16:22:23	U	SMPL
PB119247TB - 1	PB119198TB	10	-0.0245	-	7470A	PPB	26 Apr 2019 16:24:32	U	SMPL
CCV07 - 1	CCV07	23232	5.0586	-	7470A	PPB	26 Apr 2019 16:26:39	U	SMPL
CCB07 - 1	CCB07	25	-0.0212	-	7470A	PPB	26 Apr 2019 16:28:45	U	SMPL
PB119248TB - 1	PB119225TB	14	-0.0236	-	7470A	PPB	26 Apr 2019 16:30:52	U	SMPL
K2203-18LX5 - 1		49	-0.0160	-	7470A	PPB	26 Apr 2019 16:42:08	U	SMPL
K2203-18A - 1		18780	4.0841	-	7470A	PPB	26 Apr 2019 16:44:15	U	SMPL
K2545-02LX5 - 1		-76	-0.0433	-	7470A	PPB	26 Apr 2019 16:46:22	U	SMPL
K2545-02A - 1		18810	4.0907	-	7470A	PPB	26 Apr 2019 16:48:29	U	SMPL
K2198-08LX5 - 1		6	-0.0254	-	7470A	PPB	26 Apr 2019 16:58:13	U	SMPL
K2198-08A - 1		17370	3.7755	-	7470A	PPB	26 Apr 2019 17:00:20	U	SMPL
K2545-01LX5 - 1		-63	-0.0405	-	7470A	PPB	26 Apr 2019 17:02:29	U	SMPL
K2545-01A - 1		17881	3.8873	-	7470A	PPB	26 Apr 2019 17:04:37	U	SMPL
K2578-01DLX10 - 1		4741	1.0111	-	7470A	PPB	26 Apr 2019 17:06:44	U	SMPL
CCV08 - 1	CCV08	23205	5.0527	-	7470A	PPB	26 Apr 2019 17:08:51	U	SMPL
CCB08 - 1	CCB08	-37	-0.0348	-	7470A	PPB	26 Apr 2019 17:10:59	U	SMPL