



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102384

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/26/2019	15:25
2	CAL2	1	Water	NA	NA	21.1	7.00	04/26/2019	15:26
3	CAL3	1	Water	NA	NA	21.1	10.02	04/26/2019	15:27
4	ICV	1	Water	NA	NA	21.1	7.00	04/26/2019	15:29
5	CCV1	1	Water	NA	NA	21.1	2.02	04/26/2019	15:30
6	K2583-01	1	Water	NA	NA	19.2	7.94	04/26/2019	15:31
7	K2583-03	1	Water	NA	NA	19.6	5.00	04/26/2019	15:32
8	K2583-03DUP	1	Water	NA	NA	19.7	5.00	04/26/2019	15:34
9	CCV2	1	Water	NA	NA	21.2	12.02	04/26/2019	15:35

LB 102384

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2583

WorkList ID : 125380

Date : 04-26-2019 15:16:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2019	Water	K2583-01-A	pH	Cool 4 deg C	PSEG01	K11	30026-30147-30148	04/26/2019	9040C
04/26/2019	Water	K2583-03 - 1X	pH	Cool 4 deg C	PSEG01	K11	30104	04/26/2019	9040C

Date/Time 04-26-19 15:20
 Received by: Jn
 Relinquished by: Jn

Date/Time 04-26-19 16:30
 Received by: Cn
 Relinquished by: Jn