



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102385

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/26/2019	15:25
2	CAL2	1	Water	NA	NA	21.1	7.00	04/26/2019	15:26
3	CAL3	1	Water	NA	NA	21.1	10.02	04/26/2019	15:27
4	ICV	1	Water	NA	NA	21.1	7.00	04/26/2019	15:29
5	CCV1	1	Water	NA	NA	21.1	2.02	04/26/2019	15:40
6	K2198-08	1	Solid	20.03	20	23.3	7.34	04/26/2019	15:41
7	K2198-08DUP	1	Solid	20.04	20	23.4	7.35	04/26/2019	15:42
8	K2598-01	1	Solid	20.02	20	21.5	6.25	04/26/2019	15:44
9	CCV2	1	Water	NA	NA	21.2	12.02	04/26/2019	15:45

15162385

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K2599

WorkList ID : 125378

Date : 04-26-2019 15:00:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/25/2019	Solid	K2198-08	Corrosivity	Cool 4 deg C	PSEG05		OK-01-042519	04/25/2019	9045D
04/23/2019	Solid	K2598-01	Corrosivity	Cool 4 deg C	CLEA11		OIL-WATER-SEPERATOR-2(C	04/23/2019	9045D

Date/Time 04.26.19 15:00
Received by: JS
Relinquished by: JS

Date/Time 04.26.19 16:30
Received by: JS
Relinquished by: JS