



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 04/29/2019

Run Number: LB102388

BalanceID: WC SC-3

OvenID: WC-OVEN-2

FilterID: 15028098

ThermometerID: WET-OVEN 2

TEMP1 IN: 103 04/26/2019 16:00 TEMP OUT: 104 04/29/2019 07:00
TEMP2 IN: 104 04/29/2019 07:40 TEMP OUT: 104 04/29/2019 09:00
TEMP3 IN: 105 04/29/2019 10:00 TEMP OUT: 105 04/29/2019 11:45
TEMP3 IN: 105 04/29/2019 12:30 TEMP OUT: 105 04/29/2019 14:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB102388BL	LB102388BL	1.3956	1.3956	100	1.3956	1.3956	1.3956	0	0
2	LB102388BS	LB102388BS	1.3985	1.3985	100	1.449	1.449	1.449	0.0505	505
3	K2597-02	OB-722	1.409	1.409	200	1.414	1.414	1.414	0.005	25
4	K2597-02DUP	OB-722DUP	1.4014	1.4014	200	1.4063	1.4063	1.4063	0.0049	24.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS042919

WorkList ID : 125384

Date : 04-29-2019 07:04:38

13102388

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/03/2019	Water	K2597-02	TSS	Cool 4 deg C	CAMP02	K41	OB-722	04/26/2019	SM2540 D

Date/Time

Received by:

Relinquished by:

04/29/19 09:30
AS
CS

Date/Time

Received by:

Relinquished by:

04/29/19 10:20
CS
AS