



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102403

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/29/2019	09:20
2	CAL2	1	Water	NA	NA	21.1	7.00	04/29/2019	09:21
3	CAL3	1	Water	NA	NA	21.1	10.04	04/29/2019	09:22
4	ICV	1	Water	NA	NA	21.1	6.99	04/29/2019	09:24
5	CCV1	1	Water	NA	NA	21.2	2.02	04/29/2019	09:25
6	K2581-01	1	Water	NA	NA	22.3	6.49	04/29/2019	09:26
7	K2581-01DUP	1	Water	NA	NA	22.4	6.50	04/29/2019	09:27
8	K2581-02	1	Water	NA	NA	22.6	6.40	04/29/2019	09:29
9	CCV2	1	Water	NA	NA	21.2	12.02	04/29/2019	09:30

WORKLIST(Hardcopy Internal Chain)

13102403

WorkList Name : CORROSIONITY K2581

WorkList ID : 125427

Date : 04-29-2019 9:10:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/25/2019	Water	K2581-01 - E	Corrosivity	Cool 4 deg C	DVIR01		MW-03	04/25/2019	9040C
04/25/2019	Water	K2581-02 - E	Corrosivity	Cool 4 deg C	DVIR01		MW-02	04/25/2019	9040C

Date/Time 04.29.19 9:15
 Received by: JR
 Relinquished by: JT

Date/Time 04.29.19 10:20
 Received by: JT
 Relinquished by: JR