



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102404

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/29/2019	13:05
2	CAL2	1	Water	NA	NA	21.1	7.00	04/29/2019	13:06
3	CAL3	1	Water	NA	NA	21.1	10.03	04/29/2019	13:07
4	ICV	1	Water	NA	NA	21.1	7.00	04/29/2019	13:08
5	CCV1	1	Water	NA	NA	21.2	2.01	04/29/2019	13:10
6	K2609-01	1	Water	NA	NA	22.4	6.75	04/29/2019	13:11
7	CCV2	1	Water	NA	NA	21.1	12.02	04/29/2019	13:12

WORKLIST(Hardcopy Internal Chain)

LB 102404

WorkList Name : PH K2609

WorkList ID : 125425

Date : 04-29-2019 12:55:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2019	Water	K2609-01	pH	Cool 4 deg C	CHEM02	QA Office	55061-112018	04/26/2019	SM 4500-H B

Date/Time 04.29.19 13:00
 Received by: JP
 Relinquished by: JP

Date/Time 04.29.19 14:00
 Received by: JP
 Relinquished by: JP