

LB102406


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42379



A= 0.0000e+000

B= 2.3533e-004

C= -3.2304e-002

Rho= 0.9999121

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	53	0.000	53				
0.2	0.200	0.176	-0.024	884	0.0 %	884				
2.5	2.500	2.553	0.053	10984	0.0 %	10984				
5.0	5.000	4.972	-0.028	21264	0.0 %	21264				
7.5	7.500	7.579	0.079	32345	0.0 %	32345				
10.0	10.000	9.941	-0.059	42379	0.0 %	42379				

LB102406 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Apr 2019 13:34:35

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50.0	53	-	0.00007470A	PPB	29 Apr 2019 13:57:51	S	Std
0.2 - 1	50.2	884	-	0.20007470A	PPB	29 Apr 2019 13:59:59	S	Std
2.5 - 1	52.5	10984	-	2.50007470A	PPB	29 Apr 2019 14:02:07	S	Std
5.0 - 1	55.0	21264	-	5.00007470A	PPB	29 Apr 2019 14:04:15	S	Std
7.5 - 1	57.5	32345	-	7.50007470A	PPB	29 Apr 2019 14:06:22	S	Std
10.0 - 1	510.0	42379	-	10.00007470A	PPB	29 Apr 2019 14:08:29	S	Std
ICV23 - 1	ICV23	16887	3.9417	-7470A	PPB	29 Apr 2019 14:10:58	U	SMPL
ICB23 - 1	ICB23	-43	-0.0424	-7470A	PPB	29 Apr 2019 14:13:04	U	SMPL
CCV09 - 1	CCV09	20974	4.9034	-7470A	PPB	29 Apr 2019 14:15:09	U	SMPL
CCB09 - 1	CCB09	-50	-0.0441	-7470A	PPB	29 Apr 2019 14:17:15	U	SMPL
CRA - 1	CRA	852	0.1682	-7470A	PPB	29 Apr 2019 14:19:21	U	SMPL
HighStd - 1	HighStd	41961	9.8422	-7470A	PPB	29 Apr 2019 14:21:27	U	SMPL
ChkStd - 1	ChkStd	28839	6.7543	-7470A	PPB	29 Apr 2019 14:23:33	U	SMPL
PB119302BL - 1	PBW	-88	-0.0530	-7470A	PPB	29 Apr 2019 14:25:40	U	SMPL
PB119302BS - 1	LCSW	16491	3.8485	-7470A	PPB	29 Apr 2019 14:27:47	U	SMPL
K2578-02 - 1	WATER-COMP	36	-0.0238	-7470A	PPB	29 Apr 2019 14:29:54	U	SMPL
K2583-01 - 1	30026-30147-30148-30105	-10	-0.0347	-7470A	PPB	29 Apr 2019 14:32:02	U	SMPL
K2583-01DUP - 1	30026-30147-30148-30105DUP	-4	-0.0332	-7470A	PPB	29 Apr 2019 14:34:10	U	SMPL
K2583-01MS - 1	30026-30147-30148-30105MS	18609	4.3469	-7470A	PPB	29 Apr 2019 14:40:24	U	SMPL
K2583-01MSD - 1	30026-30147-30148-30105MSD	18043	4.2137	-7470A	PPB	29 Apr 2019 14:45:43	U	SMPL
CCV10 - 1	CCV10	20936	4.8945	-7470A	PPB	29 Apr 2019 14:47:49	U	SMPL
CCB10 - 1	CCB10	-74	-0.0497	-7470A	PPB	29 Apr 2019 14:49:54	U	SMPL
PB119303BL - 1	PBW	-35	-0.0405	-7470A	PPB	29 Apr 2019 14:52:01	U	SMPL
PB119303BS - 1	LCSW	16885	3.9412	-7470A	PPB	29 Apr 2019 14:54:07	U	SMPL
K2583-02 - 1	30055-23933	-3	-0.0330	-7470A	PPB	29 Apr 2019 14:56:14	U	SMPL
K2598-01 - 1	OIL-WATER-SEPERATOR-2(OWS-2)	-50	-0.0441	-7470A	PPB	29 Apr 2019 14:58:21	U	SMPL
K2598-01DUP - 1	OIL-WATER-SEPERATOR-2(OWS-2)DUP	7	-0.0307	-7470A	PPB	29 Apr 2019 15:00:27	U	SMPL
K2598-01MS - 1	OIL-WATER-SEPERATOR-2(OWS-2)MS	16615	3.8776	-7470A	PPB	29 Apr 2019 15:02:35	U	SMPL
K2598-01MSD - 1	OIL-WATER-SEPERATOR-2(OWS-2)MSD	16243	3.7901	-7470A	PPB	29 Apr 2019 15:04:42	U	SMPL
K2599-01 - 1	SB-11(6-7)	-48	-0.0436	-7470A	PPB	29 Apr 2019 15:06:50	U	SMPL
PB119304BL - 1	PBW	-20	-0.0370	-7470A	PPB	29 Apr 2019 15:08:58	U	SMPL
PB119304BS - 1	LCSW	16778	3.9160	-7470A	PPB	29 Apr 2019 15:11:04	U	SMPL
CCV11 - 1	CCV11	21022	4.9147	-7470A	PPB	29 Apr 2019 15:13:13	U	SMPL
CCB11 - 1	CCB11	-22	-0.0375	-7470A	PPB	29 Apr 2019 15:15:21	U	SMPL
K2572-01 - 1	MH-6	22	-0.0271	-7470A	PPB	29 Apr 2019 15:17:28	U	SMPL
K2581-01 - 1	MW-03	87	-0.0118	-7470A	PPB	29 Apr 2019 15:19:34	U	SMPL
K2581-02 - 1	MW-02	72	-0.0154	-7470A	PPB	29 Apr 2019 15:21:41	U	SMPL
K2581-02DUP - 1	MW-02DUP	65	-0.0170	-7470A	PPB	29 Apr 2019 15:23:47	U	SMPL
K2581-02MS - 1	MW-02MS	14157	3.2992	-7470A	PPB	29 Apr 2019 15:25:54	U	SMPL
K2581-02MSD - 1	MW-02MSD	14120	3.2905	-7470A	PPB	29 Apr 2019 15:28:01	U	SMPL
PB119302TB - 1	PB119268TB	-91	-0.0537	-7470A	PPB	29 Apr 2019 15:30:09	U	SMPL
PB119303TB - 1	PB119266TB	-19	-0.0368	-7470A	PPB	29 Apr 2019 15:32:18	U	SMPL
K2583-01LX5 - 1		-30	-0.0394	-7470A	PPB	29 Apr 2019 15:34:26	U	SMPL
K2583-01A - 1		18839	4.4010	-7470A	PPB	29 Apr 2019 15:39:52	U	SMPL
CCV12 - 1	CCV12	21363	4.9950	-7470A	PPB	29 Apr 2019 15:41:58	U	SMPL
CCB12 - 1	CCB12	-61	-0.0467	-7470A	PPB	29 Apr 2019 15:44:05	U	SMPL
K2598-01LX5 - 1		16	-0.0285	-7470A	PPB	29 Apr 2019 15:46:12	U	SMPL
K2598-01A - 1		16517	3.8546	-7470A	PPB	29 Apr 2019 15:48:19	U	SMPL
K2581-02LX5 - 1		-26	-0.0384	-7470A	PPB	29 Apr 2019 15:50:25	U	SMPL
K2581-02A - 1		13916	3.2425	-7470A	PPB	29 Apr 2019 15:52:31	U	SMPL
CCV13 - 1	CCV13	21014	4.9128	-7470A	PPB	29 Apr 2019 15:54:41	U	SMPL
CCB13 - 1	CCB13	-55	-0.0452	-7470A	PPB	29 Apr 2019 15:56:54	U	SMPL