



Analytical Summary Report

Analysis Method: 9041A

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102409

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467
0.0-3.0 pH paper refill	W1931
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
3.0-5.5 pH paper refill	W1934
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.0	04/29/2019	13:15
2	CAL2	1	Water	NA	NA	21.3	7.0	04/29/2019	13:16
3	CAL3	1	Water	NA	NA	21.4	10.0	04/29/2019	13:17
4	ICV	1	Water	NA	NA	21.4	7.0	04/29/2019	13:19
5	CCV1	1	Water	NA	NA	21.3	2.0	04/29/2019	13:20
6	K2609-02	1	Water	NA	NA	22.5	6.8	04/29/2019	13:21
7	CCV2	1	Water	NA	NA	21.4	12.0	04/29/2019	13:22

WORKLIST(Hardcopy Internal Chain)

JP 102406
1B 102409

WorkList Name : PH 042919

WorkList ID : 125426

Date : 04-29-2019 12:56:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2019	Water	K2609-02	pH	Cool 4 deg C	CHEM02	QA Office	55061-112018	04/26/2019	9041A

Date/Time 04.29.19 13:00
Received by: JO
Relinquished by: JT

Date/Time 04.29.19 14:00
Received by: JT
Relinquished by: JO