

7470A


LB102428

INSTRUMENT ID : CU1

Linear

 μ Abs.:

45495



A= 0.0000e+000

B= 2.1961e-004

C= -8.7211e-002

Rho= 0.9993182

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	202	0.000	202				
0.2	0.200	0.154	-0.046	1098	0.0 %	1098				
2.5	2.500	2.455	-0.045	11576	0.0 %	11576				
5.0	5.000	5.302	0.302	24538	0.0 %	24538				
7.5	7.500	7.428	-0.072	34220	0.0 %	34220				
10.0	10.000	9.904	-0.096	45495	0.0 %	45495				

LB102428 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Apr 2019 12:30:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	202	-	0.0000	7470A	PPB	30 Apr 2019 12:44:06	S	Std
0.2 - 1	50.2	1098	-	0.2000	7470A	PPB	30 Apr 2019 12:46:14	S	Std
2.5 - 1	52.5	11576	-	2.5000	7470A	PPB	30 Apr 2019 12:48:22	S	Std
5.0 - 1	55.0	24538	-	5.0000	7470A	PPB	30 Apr 2019 12:50:30	S	Std
7.5 - 1	57.5	34220	-	7.5000	7470A	PPB	30 Apr 2019 12:52:37	S	Std
10.0 - 1	510.0	45495	-	10.0000	7470A	PPB	30 Apr 2019 12:54:44	S	Std
ICV24 - 1	ICV24	17870	3.8373	-	7470A	PPB	30 Apr 2019 12:57:46	U	SMPL
ICB24 - 1	ICB24	134	-0.0578	-	7470A	PPB	30 Apr 2019 12:59:51	U	SMPL
CCV14 - 1	CCV14	22581	4.8719	-	7470A	PPB	30 Apr 2019 13:01:58	U	SMPL
CCB14 - 1	CCB14	280	-0.0257	-	7470A	PPB	30 Apr 2019 13:04:03	U	SMPL
CRA - 1	CRA	1082	0.1504	-	7470A	PPB	30 Apr 2019 13:06:09	U	SMPL
HighStd - 1	HighStd	44493	9.6841	-	7470A	PPB	30 Apr 2019 13:08:15	U	SMPL
ChkStd - 1	ChkStd	31667	6.8673	-	7470A	PPB	30 Apr 2019 13:10:21	U	SMPL
PB119332BL - 1	PBW	-44	-0.0969	-	7470A	PPB	30 Apr 2019 13:12:28	U	SMPL
PB119332BS - 1	LCSW	16590	3.5562	-	7470A	PPB	30 Apr 2019 13:14:35	U	SMPL
K2194-02 - 1	JC-03-042919-A	-1	-0.0874	-	7470A	PPB	30 Apr 2019 13:16:43	U	SMPL
K2194-02DUP - 1	JC-03-042919-ADUP	67	-0.0725	-	7470A	PPB	30 Apr 2019 13:18:50	U	SMPL
K2194-02MS - 1	JC-03-042919-AMS	17231	3.6970	-	7470A	PPB	30 Apr 2019 13:20:58	U	SMPL
K2194-02MSD - 1	JC-03-042919-AMSD	17022	3.6511	-	7470A	PPB	30 Apr 2019 13:23:05	U	SMPL
K2194-04 - 1	JC-03-042919-B	3	-0.0866	-	7470A	PPB	30 Apr 2019 13:25:12	U	SMPL
CCV15 - 1	CCV15	22225	4.7937	-	7470A	PPB	30 Apr 2019 13:27:18	U	SMPL
CCB15 - 1	CCB15	145	-0.0554	-	7470A	PPB	30 Apr 2019 13:29:24	U	SMPL
K2194-06 - 1	JC-03-042919-C	33	-0.0800	-	7470A	PPB	30 Apr 2019 13:31:30	U	SMPL
K2194-08 - 1	JC-03-042919-D	58	-0.0745	-	7470A	PPB	30 Apr 2019 13:33:36	U	SMPL
K2194-10 - 1	JC-03-042919-E	62	-0.0736	-	7470A	PPB	30 Apr 2019 13:35:42	U	SMPL
K2194-12 - 1	JC-03-042919-F	90	-0.0674	-	7470A	PPB	30 Apr 2019 13:37:49	U	SMPL
K2194-14 - 1	JC-03-042919-G	63	-0.0734	-	7470A	PPB	30 Apr 2019 13:39:56	U	SMPL
K2194-16 - 1	JC-03-042919-H	48	-0.0767	-	7470A	PPB	30 Apr 2019 13:42:03	U	SMPL
K2194-18 - 1	JC-03-042919-I	46	-0.0771	-	7470A	PPB	30 Apr 2019 13:44:11	U	SMPL
K2199-18 - 1	OR-02-042919-A	59	-0.0743	-	7470A	PPB	30 Apr 2019 13:46:19	U	SMPL
K2199-20 - 1	OR-02-042919-B	95	-0.0663	-	7470A	PPB	30 Apr 2019 13:48:26	U	SMPL
K2608-01 - 1	DRUM-A-E	199	-0.0435	-	7470A	PPB	30 Apr 2019 13:50:33	U	SMPL
CCV16 - 1	CCV16	22512	4.8568	-	7470A	PPB	30 Apr 2019 13:52:39	U	SMPL
CCB16 - 1	CCB16	236	-0.0354	-	7470A	PPB	30 Apr 2019 13:54:45	U	SMPL
PB119332TB - 1	PB119310TB	177	-0.0483	-	7470A	PPB	30 Apr 2019 13:56:52	U	SMPL
K2194-02LX5 - 1		291	-0.0233	-	7470A	PPB	30 Apr 2019 13:58:59	U	SMPL
K2194-02A - 1		16426	3.5202	-	7470A	PPB	30 Apr 2019 14:01:05	U	SMPL
CCV17 - 1	CCV17	22135	4.7740	-	7470A	PPB	30 Apr 2019 14:03:12	U	SMPL
CCB17 - 1	CCB17	352	-0.0099	-	7470A	PPB	30 Apr 2019 14:05:18	U	SMPL