

LB102430


7471B

INSTRUMENT ID : CU1

Linear

μ Abs.:

62599



A= 0.0000e+000

B= 1.6118e-004

C= -4.4623e-002

Rho= 0.9998782

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	242	0.000	242				
0.2	0.200	0.223	0.023	1658	0.0 %	1658				
2.5	2.500	2.560	0.060	16162	0.0 %	16162				
5.0	5.000	4.883	-0.117	30573	0.0 %	30573				
7.5	7.500	7.494	-0.006	46774	0.0 %	46774				
10.0	10.000	10.045	0.045	62599	0.0 %	62599				

LB102430 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Apr 2019 14:17:36

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	S0-0	242	-	0.00007471B	PPB	30 Apr 2019 14:18:23	S	Std
0.2 - 1	S0-2	1658	-	0.20007471B	PPB	30 Apr 2019 14:20:40	S	Std
2.5 - 1	S2-5	16162	-	2.50007471B	PPB	30 Apr 2019 14:22:59	S	Std
5.0 - 1	S5-0	30573	-	5.00007471B	PPB	30 Apr 2019 14:25:16	S	Std
7.5 - 1	S7-5	46774	-	7.50007471B	PPB	30 Apr 2019 14:27:33	S	Std
10.0 - 1	S10-0	62599	-	10.00007471B	PPB	30 Apr 2019 14:29:49	S	Std
ICV25 - 1	ICV25	25235	4.0228	-7471B	PPB	30 Apr 2019 14:32:59	U	SMPL
ICB25 - 1	ICB25	171	-0.0171	-7471B	PPB	30 Apr 2019 14:35:14	U	SMPL
CCV18 - 1	CCV18	30558	4.8807	-7471B	PPB	30 Apr 2019 14:37:31	U	SMPL
CCB18 - 1	CCB18	51	-0.0364	-7471B	PPB	30 Apr 2019 14:39:47	U	SMPL
CRA - 1	CRA	1401	0.1812	-7471B	PPB	30 Apr 2019 14:42:03	U	SMPL
HighStd - 1	HighStd	62930	10.0984	-7471B	PPB	30 Apr 2019 14:44:18	U	SMPL
ChkStd - 1	ChkStd	43266	6.9290	-7471B	PPB	30 Apr 2019 14:46:34	U	SMPL
PB119330BL - 1	PBS	-193	-0.0757	-7471B	PPB	30 Apr 2019 14:48:52	U	SMPL
PB119330BS - 1	LCSS	21733	3.4583	-7471B	PPB	30 Apr 2019 14:51:08	U	SMPL
K2194-01 - 1	JC-03-042919-A	18790	2.9840	-7471B	PPB	30 Apr 2019 14:53:25	U	SMPL
K2194-03 - 1	JC-03-042919-B	8915	1.3923	-7471B	PPB	30 Apr 2019 14:55:41	U	SMPL
K2194-05 - 1	JC-03-042919-C	10277	1.6118	-7471B	PPB	30 Apr 2019 14:57:59	U	SMPL
K2194-07 - 1	JC-03-042919-D	5191	0.7921	-7471B	PPB	30 Apr 2019 15:00:16	U	SMPL
K2194-09 - 1	JC-03-042919-E	16124	2.5542	-7471B	PPB	30 Apr 2019 15:02:31	U	SMPL
CCV19 - 1	CCV19	30519	4.8744	-7471B	PPB	30 Apr 2019 15:04:46	U	SMPL
CCB19 - 1	CCB19	126	-0.0243	-7471B	PPB	30 Apr 2019 15:07:02	U	SMPL
K2194-11 - 1	JC-03-042919-F	8590	1.3399	-7471B	PPB	30 Apr 2019 15:09:19	U	SMPL
K2194-13 - 1	JC-03-042919-G	10143	1.5902	-7471B	PPB	30 Apr 2019 15:11:34	U	SMPL
K2194-15 - 1	JC-03-042919-H	13729	2.1682	-7471B	PPB	30 Apr 2019 15:13:50	U	SMPL
K2194-17 - 1	JC-03-042919-I	9814	1.5372	-7471B	PPB	30 Apr 2019 15:16:05	U	SMPL
K2194-17DUP - 1	JC-03-042919-IDUP	9945	1.5583	-7471B	PPB	30 Apr 2019 15:18:21	U	SMPL
K2194-17MS - 1	JC-03-042919-IMS	34870	5.5757	-7471B	PPB	30 Apr 2019 15:20:38	U	SMPL
K2194-17MSD - 1	JC-03-042919-IMSD	34624	5.5361	-7471B	PPB	30 Apr 2019 15:22:54	U	SMPL
K2199-17 - 1	OR-02-042919-A	5861	0.9001	-7471B	PPB	30 Apr 2019 15:25:12	U	SMPL
K2199-19 - 1	OR-02-042919-B	1648	0.2210	-7471B	PPB	30 Apr 2019 15:27:30	U	SMPL
K2579-01 - 1	WASTE	3932	0.5891	-7471B	PPB	30 Apr 2019 15:29:45	U	SMPL
CCV20 - 1	CCV20	30598	4.8872	-7471B	PPB	30 Apr 2019 15:32:01	U	SMPL
CCB20 - 1	CCB20	109	-0.0271	-7471B	PPB	30 Apr 2019 15:34:17	U	SMPL
K2584-01 - 1	SSSI-0419-S-DH-7	7360	1.1417	-7471B	PPB	30 Apr 2019 15:36:34	U	SMPL
K2584-03 - 1	SSSI-0419-S-DH-8	6749	1.0432	-7471B	PPB	30 Apr 2019 15:38:50	U	SMPL
K2601-01 - 1	FK-EF-01-008-ABCD	142435	22.9131	-7471B	PPB	30 Apr 2019 15:41:05	U	SMPL
K2194-17LX5 - 1		1610	0.2149	-7471B	PPB	30 Apr 2019 15:43:21	U	SMPL
K2194-17A - 1		34478	5.5125	-7471B	PPB	30 Apr 2019 15:45:40	U	SMPL
K2601-01DLX5 - 1		28365	4.5272	-7471B	PPB	30 Apr 2019 15:47:56	U	SMPL
CCV21 - 1	CCV21	30623	4.8912	-7471B	PPB	30 Apr 2019 15:50:14	U	SMPL
CCB21 - 1	CCB21	110	-0.0269	-7471B	PPB	30 Apr 2019 15:52:31	U	SMPL