

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB102431BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP74163
calibration std. phosphate 1 ppm	WP74230
calibration std. phosphate 0.5 ppm	WP74229
calibration std. phosphate 0.3 ppm	WP74228
calibration std. 0 ppm	WP74225
phosphate ICV std.	WP74237
phosphate CCV std.	WP74235
calibration std. phosphate 0.1 ppm	WP74227
calibration std. phosphate 0.05 ppm	WP74226
Combined reagent	WP74247
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: 0.0002Slope: 0.611Regression: 0.999521

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/30/2019	12:30
2	CAL2	0.05	1	50	50	0.032	04/30/2019	12:30
3	CAL3	0.10	1	50	50	0.063	04/30/2019	12:31
4	CAL4	0.30	1	50	50	0.190	04/30/2019	12:31
5	CAL5	0.50	1	50	50	0.292	04/30/2019	12:33
6	CAL6	1.00	1	50	50	0.616	04/30/2019	12:33



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB102431

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.301	0.492	04/30/2019	12:34
2	ICB		1	50	50	0.000	0.000	04/30/2019	12:35
3	CCV1	0.50	1	50	50	0.299	0.489	04/30/2019	12:36
4	CCB1		1	50	50	0.000	0.000	04/30/2019	12:36
5	PB119325BL		1	1.00	50	0.000	0.000	04/30/2019	12:37
6	PB119325BS	0.50	1	1.01	50	0.344	0.563	04/30/2019	12:38
7	K2551-08		1	1.00	50	2.249	3.681	04/30/2019	12:38
8	K2551-08DUP		1	1.05	50	2.004	3.280	04/30/2019	12:40
9	K2551-08		50	1.00	50	0.616	1.008	04/30/2019	12:51
10	K2551-08DUP		50	1.05	50	0.610	0.998	04/30/2019	12:51
11	CCV2	0.50	1	50	50	0.310	0.507	04/30/2019	12:52
12	CCB2		1	50	50	0.000	0.000	04/30/2019	12:53