



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB102435

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/30/2019	12:55
2	K2581-01		1	8	100.00	>212.0	>212.0	04/30/2019	13:20
3	K2581-01DUP		1	8	100.00	>212.0	>212.0	04/30/2019	13:45
4	K2581-02		1	9	100.00	>212.0	>212.0	04/30/2019	14:10
5	K2583-01		1	9	100.00	>212.0	>212.0	04/30/2019	14:40
6	K2583-03		1	7	100.00	>212.0	>212.0	04/30/2019	15:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102435

WorkList Name : FP42919

WorkList ID : 125410

Date : 04-29-2019 09:30:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/09/2019	Water	K2581-01	E	Cool 4 deg C	DVIR01		MW-03	04/25/2019	1010A
05/09/2019	Water	K2581-02	E	Cool 4 deg C	DVIR01		MW-02	04/25/2019	1010A
05/03/2019	Water	K2583-01	C	Cool 4 deg C	PSEG01	K11	30026-30147-30148-30105	04/26/2019	1010A
05/03/2019	Water	K2583-03	A	Cool 4 deg C	PSEG01	K11	30104	04/26/2019	1010A

Date/Time 04/30/19 12:30

Received by: KS

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

04/30/19 15:15
CP
KS