

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB102436

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP74163
calibration std. phosphate 1 ppm	WP74230
calibration std. phosphate 0.5 ppm	WP74229
calibration std. phosphate 0.3 ppm	WP74228
calibration std. 0 ppm	WP74225
phosphate ICV std.	WP74237
phosphate CCV std.	WP74235
calibration std. phosphate 0.1 ppm	WP74227
calibration std. phosphate 0.05 ppm	WP74226
Combined reagent	WP74247
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: 0.0002Slope: 0.611Regression: 0.999521

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/30/2019	12:30
2	CAL2	0.05	1	50	50	0.032	04/30/2019	12:30
3	CAL3	0.10	1	50	50	0.063	04/30/2019	12:31
4	CAL4	0.30	1	50	50	0.190	04/30/2019	12:31
5	CAL5	0.50	1	50	50	0.292	04/30/2019	12:33
6	CAL6	1.00	1	50	50	0.616	04/30/2019	12:33



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB102436

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.301	0.492	04/30/2019	12:34
2	ICB		1	50	50	0.000	0.000	04/30/2019	12:35
3	CCV1	0.50	1	50	50	0.310	0.507	04/30/2019	12:52
4	CCB1		1	50	50	0.000	0.000	04/30/2019	12:53
5	PB119324BL		1	50	50	0.000	0.000	04/30/2019	12:54
6	PB119324BS	0.50	1	50	50	0.331	0.541	04/30/2019	12:54
7	K2597-02		1	50	50	0.214	0.350	04/30/2019	12:55
8	K2597-08		1	50	50	0.149	0.244	04/30/2019	12:55
9	K2597-08DUP		1	50	50	0.148	0.242	04/30/2019	12:56
10	K2597-08MS	0.50	1	50	50	0.439	0.718	04/30/2019	12:58
11	K2597-08MSD	0.50	1	50	50	0.440	0.720	04/30/2019	12:58
12	CCV2	0.50	1	50	50	0.312	0.510	04/30/2019	12:59
13	CCB2		1	50	50	0.000	0.000	04/30/2019	13:00