



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102446

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/30/2019	17:30
2	CAL2	1	Water	NA	NA	21.1	7.01	04/30/2019	17:30
3	CAL3	1	Water	NA	NA	21.1	10.02	04/30/2019	17:31
4	ICV	1	Water	NA	NA	21.1	6.99	04/30/2019	17:33
5	CCV1	1	Water	NA	NA	21.2	2.02	04/30/2019	17:35
6	K2191-04	1	Solid	20.02	20	22.4	6.91	04/30/2019	17:36
7	K2191-04DUP	1	Solid	20.03	20	22.5	6.92	04/30/2019	17:38
8	K2204-06	1	Solid	20.05	20	22.5	9.51	04/30/2019	17:39
9	K2617-03	1	Solid	20.04	20	22.0	9.22	04/30/2019	17:40
10	CCV2	1	Water	NA	NA	21.2	12.01	04/30/2019	17:41

LB 102446

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2191

WorkList ID : 125478

Date : 04-30-2019 16:20:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2019	Solid	K2191-04	Corrosivity	Cool 4 deg C	PSEG01	K31	P-66-TP	04/30/2019	9045D
04/30/2019	Solid	K2204-06	Corrosivity	Cool 4 deg C	PSEG05		TR-04-043019	04/30/2019	9045D
04/30/2019	Solid	K2617-03	Corrosivity	Cool 4 deg C	PSEG01	K21	ROCK-COMP	04/30/2019	9045D

Date/Time 04.30.19 16:25
 Received by: Jg
 Relinquished by: cd

Date/Time 04.30.19 18:00
 Received by: [Signature]
 Relinquished by: Jn