



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102447

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/30/2019	17:30
2	CAL2	1	Water	NA	NA	21.1	7.01	04/30/2019	17:31
3	CAL3	1	Water	NA	NA	21.1	10.02	04/30/2019	17:32
4	ICV	1	Water	NA	NA	21.1	6.99	04/30/2019	17:33
5	CCV1	1	Water	NA	NA	21.1	2.02	04/30/2019	17:45
6	K2623-01	1	Solid	20.02	20	22.3	8.41	04/30/2019	17:46
7	K2623-01DUP	1	Solid	20.03	20	22.4	8.42	04/30/2019	17:48
8	CCV2	1	Water	NA	NA	21.1	12.01	04/30/2019	17:50

19 102448

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K2623

WorkList ID : 125481

Date : 04-30-2019 16:22:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/29/2019	Solid	K2623-01	☒ pH	Cool 4 deg C	AREC01	K41	SP-1	04/29/2019	9045D

Date/Time 04.30.19 16:25
Received by: JS
Relinquished by: CS

Date/Time 04.30.19 18:00
Received by: JS
Relinquished by: CS