



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102448

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/30/2019	17:30
2	CAL2	1	Water	NA	NA	21.1	7.01	04/30/2019	17:31
3	CAL3	1	Water	NA	NA	21.1	10.02	04/30/2019	17:32
4	ICV	1	Water	NA	NA	21.1	6.99	04/30/2019	17:33
5	CCV1	1	Water	NA	NA	21.2	2.02	04/30/2019	18:00
6	K2617-05	1	Water	NA	NA	17.2	7.70	04/30/2019	18:01
7	K2621-02	1	Water	NA	NA	17.6	8.65	04/30/2019	18:03
8	K2621-02DUP	1	Water	NA	NA	17.7	8.66	04/30/2019	18:05
9	CCV2	1	Water	NA	NA	21.2	12.01	04/30/2019	18:06

19102448

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2621

WorkList ID : 125480

Date : 04-30-2019 16:22:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2019	Water	K2617-05- ☒	Corrosivity	Cool 4 deg C	PSEG01	K21	SUMP-PIT	04/30/2019	9040C
04/30/2019	Water	K2621-02- ☒	Corrosivity	Cool 4 deg C	PSEG01	K11	A5327	04/30/2019	9040C

Date/Time 04.30.19 16:25
 Received by: Sn
 Relinquished by: CS

Date/Time 04.30.19 18:00
 Received by: CS
 Relinquished by: 38