



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102453

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/30/2019	17:30
2	CAL2	1	Water	NA	NA	21.1	7.01	04/30/2019	17:31
3	CAL3	1	Water	NA	NA	21.1	10.02	04/30/2019	17:32
4	ICV	1	Water	NA	NA	21.1	6.99	04/30/2019	17:33
5	CCV1	1	Water	NA	NA	97.8	2.02	04/30/2019	18:00
6	K2617-04	1	Water	NA	NA	17.2	7.70	04/30/2019	18:01
7	K2621-01	1	Water	NA	NA	17.6	8.65	04/30/2019	18:03
8	K2621-01DUP	1	Water	NA	NA	17.7	8.66	04/30/2019	18:05
9	K2622-01	1	Water	NA	NA	16.5	4.48	04/30/2019	18:07
10	K2622-02	1	Water	NA	NA	19.3	4.80	04/30/2019	18:10
11	K2622-03	1	Water	NA	NA	19.9	4.91	04/30/2019	18:11
12	K2622-04	1	Water	NA	NA	19.5	5.04	04/30/2019	18:12
13	CCV2	1	Water	NA	NA	21.2	12.01	04/30/2019	18:15

LB 102453

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2621

WorkList ID : 125479

Date : 04-30-2019 16:21:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2019	Water	K2617-04 -17	pH	Cool 4 deg C	PSEG01	K21	SUMP-PIT	04/30/2019	9040C
04/30/2019	Water	K2621-01 -17	pH	Cool 4 deg C	PSEG01	K11	A5327	04/30/2019	9040C
04/30/2019	Water	K2622-01 -17	pH	Cool 4 deg C	NYUL01	K51	TH-4-30-DI-1	04/30/2019	9040C
04/30/2019	Water	K2622-02 -17	pH	Cool 4 deg C	NYUL01	K51	TH-4-30-DI-2	04/30/2019	9040C
04/30/2019	Water	K2622-03 -17	pH	Cool 4 deg C	NYUL01	K51	TH-4-30-DI-3	04/30/2019	9040C
04/30/2019	Water	K2622-04 -17	pH	Cool 4 deg C	NYUL01	K51	TH-4-04-DI-4	04/30/2019	9040C

Date/Time 04-30-19 16:25

Received by: JARelinquished by: CA

Date/Time 04-30-19 18:20

Received by: CSRelinquished by: JP