



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102469

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/01/2019	12:05
2	CAL2	1	Water	NA	NA	21.1	7.01	05/01/2019	12:06
3	CAL3	1	Water	NA	NA	21.1	10.03	05/01/2019	12:07
4	ICV	1	Water	NA	NA	21.1	7.00	05/01/2019	12:08
5	CCV1	1	Water	NA	NA	21.1	2.01	05/01/2019	12:10
6	K2629-01	1	Water	NA	NA	21.8	7.72	05/01/2019	12:12
7	K2629-01DUP	1	Water	NA	NA	21.9	7.73	05/01/2019	12:14
8	K2630-01	1	Water	NA	NA	21.8	7.38	05/01/2019	12:15
9	K2632-01	1	Water	NA	NA	21.2	4.40	05/01/2019	12:16
10	CCV2	1	Water	NA	NA	21.2	12.02	05/01/2019	12:19

LB 102469

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2629

WorkList ID : 125517

Date : 05-01-2019 11:51:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2019	Water	K2629-01 - F	pH	Cool 4 deg C	MINO01	K41	020-OUTFALL	04/30/2019	9040C
04/30/2019	Water	K2630-01 - F	pH	Cool 4 deg C	MINO01	K51	025-OUTFALL	04/30/2019	9040C
04/29/2019	Water	K2632-01 - F	pH	Cool 4 deg C	STER14	K11	MINI-CELL	04/29/2019	9040C

Date/Time 05-01-19 12:00
Received by: Jy
Relinquished by: cd

Date/Time 05-01-19 13:00
Received by: Jy
Relinquished by: Jy