



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102476

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/01/2019	15:00
2	CAL2	1	Water	NA	NA	21.1	6.99	05/01/2019	15:01
3	CAL3	1	Water	NA	NA	21.1	10.03	05/01/2019	15:02
4	ICV	1	Water	NA	NA	21.1	7.01	05/01/2019	15:05
5	CCV1	1	Water	NA	NA	21.1	2.02	05/01/2019	15:06
6	K2650-02	1	Water	NA	NA	20.2	7.16	05/01/2019	15:07
7	K2650-02DUP	1	Water	NA	NA	20.3	7.17	05/01/2019	15:09
8	K2653-02	1	Water	NA	NA	19.9	4.87	05/01/2019	15:10
9	CCV2	1	Water	NA	NA	21.1	12.02	05/01/2019	15:11

LB 102476

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2650 WorkList ID : 125536 Date : 05-01-2019 14:42:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/01/2019	Water	K2650-02 - ☒	Corrosivity	Cool 4 deg C	PSEG01	K21	GROUND-WATER	05/01/2019	9040C
05/01/2019	Water	K2653-02 - ☒	Corrosivity	Cool 4 deg C	PSEG01	K11	MR-SUMP-PIT	05/01/2019	9040C

Date/Time 05-01-19 14:50
Received by: g n
Relinquished by: CS

Date/Time 05-01-19 15:35
Received by: as
Relinquished by: ajp