



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB102481

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	05/01/2019	11:10
2	K2617-05		1	8	100.00	>212.0	>212.0	05/01/2019	11:40
3	K2617-05DUP		1	8	100.00	>212.0	>212.0	05/01/2019	12:10
4	K2621-02		1	9	100.00	>212.0	>212.0	05/01/2019	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102481

WorkList Name : FP5119

WorkList ID : 125506

Date : 05-01-2019 09:18:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/03/2019	Water	K2617-05	Flash Point	Cool 4 deg C	PSEG01	K21	SUMP-PJT	04/30/2019	1010A
05/03/2019	Water	K2621-02	Flash Point	Cool 4 deg C	PSEG01	K11	A5327	04/30/2019	1010A

Date/Time 05/01/19 10:45
 Received by: KS
 Relinquished by: CP

Date/Time 05/01/19 12:50
 Received by: CP
 Relinquished by: KS