



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102485

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.0	4.01	05/01/2019	17:00
2	CAL2	1	Water	NA	NA	21.0	7.01	05/01/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	05/01/2019	17:02
4	ICV	1	Water	NA	NA	21.0	7.00	05/01/2019	17:04
5	CCV1	1	Water	NA	NA	21.1	2.02	05/01/2019	17:05
6	K2627-02	1	Solid	20.03	20	21.7	7.77	05/01/2019	17:06
7	K2627-02DUP	1	Solid	20.04	20	21.6	7.78	05/01/2019	17:08
8	K2650-04	1	Solid	20.02	20	21.7	5.61	05/01/2019	17:10
9	K2652-02	1	Solid	20.03	20	21.6	9.27	05/01/2019	17:11
10	K2654-04	1	Solid	20.05	20	21.9	5.51	05/01/2019	17:12
11	CCV2	1	Water	NA	NA	21.1	12.02	05/01/2019	17:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2627

WorkList ID : 125543

Date : 05-01-2019 15:23:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2019	Solid	K2627-02	Corrosivity	Cool 4 deg C	AREC01	K43	WC-1	04/30/2019	9045D
05/01/2019	Solid	K2650-04	Corrosivity	Cool 4 deg C	PSEG01	K21	OILY-DEBRIS	05/01/2019	9045D
05/01/2019	Solid	K2652-02	Corrosivity	Cool 4 deg C	PSEG01	K11	DRUM-1-4	05/01/2019	9045D
05/01/2019	Solid	K2654-04	Corrosivity	Cool 4 deg C	PSEG01	K31	OILY-STONES	05/01/2019	9045D

Date/Time 05-01-19 15:25
 Received by: JA
 Relinquished by: CA

Date/Time 05-01-19 17:00
 Received by: JA
 Relinquished by: JA