



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB102489
Analysis Date: 05/02/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 05/02/2019
Extraction IN Time: 07:25
Extraction OUT Time: 07:50

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB102489BL	LB102489BL	WATER	1.3	1000	100	2.4106	2.4106	3.02	2.4106	2.4106	0.0000	0
2	LB102489BS	LB102489BS	WATER	1.3	1000	100	2.7468	2.7468	3.01	2.7659	2.7659	0.0191	19.1
3	LB102489BSD	LB102489BSD	WATER	1.3	1000	100	2.1036	2.1036	3.03	2.1229	2.1229	0.0193	19.3
4	K2650-01	GROUND-WATER	WATER	1.3	1000	100	3.1160	3.1160	3.04	3.1251	3.1251	0.0091	9.1
5	K2653-01	MR-SUMP-PIT	WATER	1.3	1000	100	3.0403	3.0403	3.03	3.0407	3.0407	0.0004	0.4
6	K2655-02	GRAB	WATER	1.6	800	100	3.1294	3.1294	3.05	3.1424	3.1424	0.0130	16.25



QC Batch# LB102489 Test: TPH Analysis Date: 05/02/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2515
pH Paper 0-14	M4240
Sodium Sulfate	WP1944
1:1 HCL	WP74116
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	10.00 ML	WP74118
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:26</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>08:25</u>		
Bal Check Time:	<u>07:35</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>09:50</u>
		Out Time:	<u>09:25</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>10:56</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>10:20</u>		
Bal Check Time:	<u>11:50</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:25</u>
		Out Time:	<u>10:55</u>		

LD102489

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K2655

WorkList ID : 125548

Date : 05-02-2019 7:15:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2019	Water	K2650-01 - K	TPH	Conc H2SO4 4c	PSEG01	K21	GROUND-WATER	05/01/2019	1664A
05/07/2019	Water	K2653-01 - N	TPH	Conc H2SO4 4c	PSEG01	K11	MR-SUMP-PIT	05/01/2019	1664A
05/08/2019	Water	K2655-02 ✓	TPH	Conc H2SO4 4c	ARAM01	K41	GRAB	05/01/2019	1664A

Date/Time 05/02/19 7:20

Received by: Jo

Relinquished by: CA

Date/Time 05/02/19 9:35

Received by: [Signature]

Relinquished by: Jo