

LB102507


7471B

INSTRUMENT ID: C01

Linear ▼

 μ Abs.:

62085



A= 0.0000e+000

B= 1.5937e-004

C= -2.1337e-002

Rho= 0.9997314

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	60	0.000	60				
0.2	0.200	0.184	-0.016	1288	0.0 %	1288				
2.5	2.500	2.439	-0.061	15436	0.0 %	15436				
5.0	5.000	5.096	0.096	32107	0.0 %	32107				
7.5	7.500	7.620	0.120	47950	0.0 %	47950				
10.0	10.000	9.873	-0.127	62085	0.0 %	62085				

LB102507 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 May 2019 14:17:36

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	S0-0	60	-	0.00007471B	PPB	02 May 2019 14:18:57	S	Std
0.2 - 1	S0-2	1288	-	0.20007471B	PPB	02 May 2019 14:21:14	S	Std
2.5 - 1	S2-5	15436	-	2.50007471B	PPB	02 May 2019 14:23:32	S	Std
5.0 - 1	S5-0	32107	-	5.00007471B	PPB	02 May 2019 14:25:49	S	Std
7.5 - 1	S7-5	47950	-	7.50007471B	PPB	02 May 2019 14:28:06	S	Std
10.0 - 1	S10-0	62085	-	10.00007471B	PPB	02 May 2019 14:38:11	S	Std
ICV27 - 1	ICV27	24306	3.8523	-7471B	PPB	02 May 2019 14:42:11	U	SMPL
ICB27 - 1	ICB27	-42	-0.0280	-7471B	PPB	02 May 2019 14:44:26	U	SMPL
CCV27 - 1	CCV27	31876	5.0587	-7471B	PPB	02 May 2019 14:46:42	U	SMPL
CCB27 - 1	CCB27	-49	-0.0291	-7471B	PPB	02 May 2019 14:48:57	U	SMPL
CRA - 1	CRA	1271	0.1812	-7471B	PPB	02 May 2019 14:51:13	U	SMPL
HighStd - 1	HighStd	62346	9.9147	-7471B	PPB	02 May 2019 14:53:29	U	SMPL
ChkStd - 1	ChkStd	44282	7.0359	-7471B	PPB	02 May 2019 14:55:45	U	SMPL
PB119380BL - 1	PBS	-105	-0.0381	-7471B	PPB	02 May 2019 14:58:06	U	SMPL
PB119380BS - 1	LCSS	20923	3.3132	-7471B	PPB	02 May 2019 15:00:24	U	SMPL
K2191-01 - 1	P-66-TP	74842	11.9062	-7471B	PPB	02 May 2019 15:02:42	U	SMPL
K2204-05 - 1	TR-04-043019	3265	0.4990	-7471B	PPB	02 May 2019 15:04:58	U	SMPL
K2617-01 - 1	ROCK-COMP	9633	1.5139	-7471B	PPB	02 May 2019 15:07:17	U	SMPL
K2623-01 - 1	SP-1	12905	2.0353	-7471B	PPB	02 May 2019 15:09:35	U	SMPL
K2623-01DUP - 1	SP-1DUP	12554	1.9794	-7471B	PPB	02 May 2019 15:11:50	U	SMPL
CCV28 - 1	CCV28	32161	5.1042	-7471B	PPB	02 May 2019 15:14:06	U	SMPL
CCB28 - 1	CCB28	-45	-0.0285	-7471B	PPB	02 May 2019 15:16:23	U	SMPL
K2623-01MS - 1	SP-1MS	37467	5.9498	-7471B	PPB	02 May 2019 15:18:47	U	SMPL
K2623-01MSD - 1	SP-1MSD	37578	5.9675	-7471B	PPB	02 May 2019 15:21:05	U	SMPL
K2627-01 - 1	WC-1	3602	0.5527	-7471B	PPB	02 May 2019 15:23:21	U	SMPL
K2652-01 - 1	DRUM-1-4	610	0.0759	-7471B	PPB	02 May 2019 15:25:38	U	SMPL
K2654-01 - 1	OILY-STONES	364	0.0367	-7471B	PPB	02 May 2019 15:27:55	U	SMPL
K2623-01LX5 - 1		2558	0.3863	-7471B	PPB	02 May 2019 15:30:12	U	SMPL
K2623-01A - 1		37165	5.9016	-7471B	PPB	02 May 2019 15:32:27	U	SMPL
K2191-01DLX2 - 1		36992	5.8741	-7471B	PPB	02 May 2019 15:34:45	U	SMPL
CCV29 - 1	CCV29	32145	5.1016	-7471B	PPB	02 May 2019 15:37:03	U	SMPL
CCB29 - 1	CCB29	77	-0.0091	-7471B	PPB	02 May 2019 15:39:20	U	SMPL