



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102513

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP74367
Residual chlorine std, 0PPM	WP74370
Residual chlorine std, 0.1PPM	WP74364
Residual chlorine std, 0.2PPM	WP74365
Residual chlorine std, and CCV , 0.4PPM	WP74366
Residual chlorine std, 1.6PPM	WP74368
Residual chlorine ICV-LCS, 0.4PPM	WP74369
Total Chlorine Powder Pillows	W2249

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	05/02/2019	09:25
2	CAL2	0.1	1	0.000	0.100	0.100	05/02/2019	09:26
3	CAL3	0.2	1	0.000	0.200	0.200	05/02/2019	09:27
4	CAL4	0.4	1	0.000	0.400	0.400	05/02/2019	09:28
5	CAL5	0.8	1	0.000	0.800	0.800	05/02/2019	09:29
6	CAL6	1.6	1	0.000	1.600	1.600	05/02/2019	09:30



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102513

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3900	05/02/2019	09:31
2	ICB				1	0.0000	0.0000	0.0000	0.0000	05/02/2019	09:32
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.4000	05/02/2019	09:32
4	CCB1				1	0.0000	0.0000	0.0000	0.0000	05/02/2019	09:33
5	LB102513BL	50	50		1	0.0000	0.0000	0.0000	0.0000	05/02/2019	09:34
6	LB102513BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.4000	05/02/2019	09:35
7	K2675-01	50	50		1	0.0000	0.0000	0.0000	0.0000	05/02/2019	09:36
8	K2675-01DUP	50	50		1	0.0000	0.0000	0.0000	0.0000	05/02/2019	09:37
9	K2675-01MS	50	50	0.4	1	0.0000	0.3700	0.3700	0.3700	05/02/2019	09:39
10	K2675-01MSD	50	50	0.4	1	0.0000	0.3800	0.3800	0.3800	05/02/2019	09:40
11	CCV2			0.4	1	0.0000	0.4000	0.4000	0.4000	05/02/2019	09:41
12	CCB2				1	0.0000	0.0000	0.0000	0.0000	05/02/2019	09:42

CB102813

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCL050219

WorkList ID : 125593

Date : 05-02-2019 09:17:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/01/2019	Water	K2675-01	Residual Chlorine	Cool 4 deg C	SHAW22	K41	OUTFALL001-WK-1	05/01/2019	SM4500 Cl G

Date/Time 05-02-19 9:20
Received by: ne
Relinquished by: CG

Date/Time 05-02-19 10:12
Received by: CG
Relinquished by: ne