



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102514

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/02/2019	11:40
2	K2650-02		1	8	100.00	>212.0	>212.0	05/02/2019	12:15
3	K2650-02DUP		1	8	100.00	>212.0	>212.0	05/02/2019	12:50
4	K2653-02		1	9	100.00	>212.0	>212.0	05/02/2019	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102574

WorkList Name : FP5219

WorkList ID : 125578

Date : 05-02-2019 09:27:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2019	Water	K2650-02	C Flash Point	Cool 4 deg C	PSEG01	K21	GROUND-WATER	05/01/2019	1010A
05/07/2019	Water	K2653-02	C Flash Point	Cool 4 deg C	PSEG01	K11	MR-SUMP-PIT	05/01/2019	1010A

Date/Time 05/02/19 11:15
 Received by: KS
 Relinquished by: CP

Date/Time 05/02/19 13:25
 Received by: CP
 Relinquished by: KS