

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102519

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/02/2019	18:15
2	CAL2	1	Water	NA	NA	21.1	7.01	05/02/2019	18:16
3	CAL3	1	Water	NA	NA	21.1	10.03	05/02/2019	18:17
4	ICV	1	Water	NA	NA	21.1	7.00	05/02/2019	18:20
5	CCV1	1	Water	NA	NA	21.1	2.01	05/02/2019	18:21
6	K2635-02	1	Solid	20.03	20	23.9	9.05	05/02/2019	18:22
7	K2635-02DUP	1	Solid	20.04	20	23.8	9.06	05/02/2019	18:24
8	K2635-04	1	Solid	20.03	20	24.4	9.07	05/02/2019	18:25
9	K2635-06	1	Solid	20.02	20	24.4	7.74	05/02/2019	18:26
10	K2635-08	1	Solid	20.01	20	24.4	10.42	05/02/2019	18:29
11	K2642-02	1	Solid	20.03	20	24.6	8.00	05/02/2019	18:30
12	K2647-02	1	Solid	20.04	20	24.8	8.36	05/02/2019	18:31
13	K2647-04	1	Solid	20.02	20	24.7	8.31	05/02/2019	18:32
14	K2668-02	1	Solid	20.03	20	23.9	6.60	05/02/2019	18:34
15	K2668-04	1	Solid	20.02	20	23.4	5.86	05/02/2019	18:35
16	CCV2	1	Water	NA	NA	21.2	12.02	05/02/2019	18:36
17	K2668-06	1	Solid	20.01	20	23.5	5.88	05/02/2019	18:37
18	K2674-01	1	Solid	20.03	20	23.9	8.66	05/02/2019	18:39
19	K2674-02	1	Solid	20.04	20	23.9	8.66	05/02/2019	18:37
20	CCV3	1	Water	NA	NA	21.2	2.02	05/02/2019	18:39

LB 102519

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2680

WorkList ID : 125599

Date : 05-02-2019 15:53:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2019	Solid	K2635-02	Corrosivity	Cool 4 deg C	PSEG05	K31	AU-05-050219-A	05/02/2019	9045D
05/02/2019	Solid	K2635-04	Corrosivity	Cool 4 deg C	PSEG05	K31	AU-05-050219-B	05/02/2019	9045D
05/02/2019	Solid	K2635-06	Corrosivity	Cool 4 deg C	PSEG05	K31	AU-05-050219-C	05/02/2019	9045D
05/02/2019	Solid	K2635-08	Corrosivity	Cool 4 deg C	PSEG05	K31	AU-05-050219-D	05/02/2019	9045D
05/02/2019	Solid	K2642-02	Corrosivity	Cool 4 deg C	PSEG05	K11	OK-02-050219	05/02/2019	9045D
05/02/2019	Solid	K2647-02	Corrosivity	Cool 4 deg C	PSEG05	K21	SU-04-050219-A	05/02/2019	9045D
05/02/2019	Solid	K2647-04	Corrosivity	Cool 4 deg C	PSEG05	K21	SU-04-050219-B	05/02/2019	9045D
05/02/2019	Solid	K2668-02	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-1	05/02/2019	9045D
05/02/2019	Solid	K2668-04	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-2	05/02/2019	9045D
05/02/2019	Solid	K2668-06	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-3	05/02/2019	9045D
05/02/2019	Solid	K2674-01	Corrosivity	Cool 4 deg C	PSEG01	K11	FILTERBAG-RO	05/02/2019	9045D
05/02/2019	Solid	K2674-02	Corrosivity	Cool 4 deg C	PSEG01	K11	FILTERBAG-RO	05/02/2019	9045D

Date/Time 05.02.19 16:20
 Received by: Jn
 Relinquished by: Jn

Date/Time 05.02.19 18:35
 Received by: Jn
 Relinquished by: Jn