

LB102533


7471B

INSTRUMENT ID: C01

Linear

 μ Abs.:

61802



A= 0.0000e+000

B= 1.6075e-004

C= 5.6626e-003

Rho= 0.9998126

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.049	0.049	269	0.000	269				
0.2	0.200	0.208	0.008	1261	0.0 %	1261				
2.5	2.500	2.384	-0.116	14792	0.0 %	14792				
5.0	5.000	5.015	0.015	31159	0.0 %	31159				
7.5	7.500	7.604	0.104	47268	0.0 %	47268				
10.0	10.000	9.941	-0.059	61802	0.0 %	61802				

LB102533 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 May 2019 10:39:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0-0	269	-	0.00007471B	PPB		03 May 2019 10:39:30	S	Std
0.2 - 1	S0-2	1261	-	0.20007471B	PPB		03 May 2019 10:41:47	S	Std
2.5 - 1	S2-5	14792	-	2.50007471B	PPB		03 May 2019 10:44:06	S	Std
5.0 - 1	S5-0	31159	-	5.00007471B	PPB		03 May 2019 10:46:23	S	Std
7.5 - 1	S7-5	47268	-	7.50007471B	PPB		03 May 2019 10:48:39	S	Std
10.0 - 1	S10-0	61802	-	10.00007471B	PPB		03 May 2019 10:50:56	S	Std
ICV28 - 1	ICV28	24518	3.9470	-7471B	PPB		03 May 2019 10:54:13	U	SMPL
ICB28 - 1	ICB28	29	0.0103	-7471B	PPB		03 May 2019 10:56:30	U	SMPL
CCV30 - 1	CCV30	31352	5.0456	-7471B	PPB		03 May 2019 10:58:45	U	SMPL
CCB30 - 1	CCB30	17	0.0084	-7471B	PPB		03 May 2019 11:01:00	U	SMPL
CRA - 1	CRA	1280	0.2114	-7471B	PPB		03 May 2019 11:03:16	U	SMPL
HighStd - 1	HighStd	62575	10.0648	-7471B	PPB		03 May 2019 11:05:31	U	SMPL
ChkStd - 1	ChkStd	44746	7.1987	-7471B	PPB		03 May 2019 11:07:46	U	SMPL
PB119441BL - 1	PBS	-15	0.0033	-7471B	PPB		03 May 2019 11:10:04	U	SMPL
PB119441BS - 1	LCSS	22510	3.6242	-7471B	PPB		03 May 2019 11:12:21	U	SMPL
K2635-01 - 1	AU-05-050219-A	4429	0.7176	-7471B	PPB		03 May 2019 11:14:38	U	SMPL
K2635-03 - 1	AU-05-050219-B	3386	0.5500	-7471B	PPB		03 May 2019 11:16:55	U	SMPL
K2635-05 - 1	AU-05-050219-C	4409	0.7144	-7471B	PPB		03 May 2019 11:19:12	U	SMPL
K2635-07 - 1	AU-05-050219-D	9477	1.5291	-7471B	PPB		03 May 2019 11:21:29	U	SMPL
K2642-01 - 1	OK-02-050219	2596	0.4230	-7471B	PPB		03 May 2019 11:23:44	U	SMPL
CCV31 - 1	CCV31	32096	5.1652	-7471B	PPB		03 May 2019 11:26:00	U	SMPL
CCB31 - 1	CCB31	0	0.0057	-7471B	PPB		03 May 2019 11:28:15	U	SMPL
K2647-01 - 1	SU-04-050219-A	1803	0.2955	-7471B	PPB		03 May 2019 11:30:31	U	SMPL
K2647-03 - 1	SU-04-050219-B	1112	0.1844	-7471B	PPB		03 May 2019 11:32:46	U	SMPL
K2667-01 - 1	SP-1-A	476	0.0822	-7471B	PPB		03 May 2019 11:35:01	U	SMPL
K2667-03 - 1	SP-1-B	1895	0.3103	-7471B	PPB		03 May 2019 11:37:17	U	SMPL
K2668-01 - 1	TP-1	644	0.1092	-7471B	PPB		03 May 2019 11:39:33	U	SMPL
K2668-03 - 1	TP-2	1174	0.1944	-7471B	PPB		03 May 2019 11:41:49	U	SMPL
K2668-05 - 1	TP-3	1359	0.2241	-7471B	PPB		03 May 2019 11:44:07	U	SMPL
K2677-01 - 1	FK-EF-01-008-ABCD	3907	0.6337	-7471B	PPB		03 May 2019 11:46:24	U	SMPL
K2678-01 - 1	FK-0D015-03-003	1406	0.2317	-7471B	PPB		03 May 2019 11:48:41	U	SMPL
K2678-01DUP - 1	FK-0D015-03-003DUP	1366	0.2253	-7471B	PPB		03 May 2019 11:50:57	U	SMPL
CCV32 - 1	CCV32	32212	5.1838	-7471B	PPB		03 May 2019 11:53:12	U	SMPL
CCB32 - 1	CCB32	-119	-0.0135	-7471B	PPB		03 May 2019 11:55:28	U	SMPL
K2678-01MS - 1	FK-0D015-03-003MS	22656	3.6477	-7471B	PPB		03 May 2019 11:57:46	U	SMPL
K2678-01MSD - 1	FK-0D015-03-003MSD	22541	3.6292	-7471B	PPB		03 May 2019 12:00:03	U	SMPL
K2678-01LX5 - 1		450	0.0780	-7471B	PPB		03 May 2019 12:02:19	U	SMPL
K2678-01A - 1		22789	3.6691	-7471B	PPB		03 May 2019 12:04:35	U	SMPL
CCV33 - 1	CCV33	30506	4.9096	-7471B	PPB		03 May 2019 12:06:50	U	SMPL
CCB33 - 1	CCB33	-42	-0.0011	-7471B	PPB		03 May 2019 12:09:07	U	SMPL