



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102548

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/03/2019	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	05/03/2019	16:11
3	CAL3	1	Water	NA	NA	20.1	10.02	05/03/2019	16:12
4	ICV	1	Water	NA	NA	20.1	6.99	05/03/2019	16:13
5	CCV1	1	Water	NA	NA	20.1	2.02	05/03/2019	16:15
6	K2674-04	1	Water	NA	NA	22.4	7.56	05/03/2019	16:16
7	K2696-02	1	Water	NA	NA	19.8	6.67	05/03/2019	16:17
8	K2696-02DUP	1	Water	NA	NA	19.7	6.68	05/03/2019	16:19
9	CCV2	1	Water	NA	NA	20.2	12.01	05/03/2019	16:20

13 102548

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2696

WorkList ID : 125659

Date : 05-03-2019 16:02:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2019	Water	K2674-04 -A	Corrosivity	Cool 4 deg C	PSEG01	K11	SV28545L	05/02/2019	9040C
05/02/2019	Water	K2696-02 -A	Corrosivity	Cool 4 deg C	PARS02	K31	WC-20190502	05/02/2019	9040C

Date/Time 05-03-19 16:00
 Received by: Jn
 Relinquished by: BP

Date/Time 05-03-19 17:00
 Received by: BP
 Relinquished by: Jn