



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102549

Slope : 98.1

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2445         |
| buffer solution pH 7 yellow    | W2358         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2391         |
| buffer solution pH 7 yellow    | W2446         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2467         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 05/03/2019 | 16:10     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.01        | 05/03/2019 | 16:11     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.02       | 05/03/2019 | 16:12     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 6.99        | 05/03/2019 | 16:13     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.02        | 05/03/2019 | 16:15     |
| 6   | K2674-03    | 1  | Water  | NA          | NA          | 22.4             | 7.56        | 05/03/2019 | 16:16     |
| 7   | K2674-03DUP | 1  | Water  | NA          | NA          | 22.5             | 7.57        | 05/03/2019 | 16:17     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.01       | 05/03/2019 | 16:20     |

28 102549

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2696      WorkList ID : 125660      Date : 05-03-2019 16:02:42

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|-------------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 05/02/2019 | Water  | K2674-03-18 | pH   | Cool 4 deg C | PSEG01   | K11              | SV28545L        | 05/02/2019   | 9040C  |

Date/Time 05.03.19 16:00  
Received by: JO  
Relinquished by: JO

Date/Time 05.03.19 17:00  
Received by: JO  
Relinquished by: JO