



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102553

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/03/2019	15:05
2	K2674-04		1	8	100.00	>212.0	>212.0	05/03/2019	15:35
3	K2674-04DUP		1	8	100.00	>212.0	>212.0	05/03/2019	15:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102553

WorkList Name : FP5319

WorkList ID : 125639

Date : 05-03-2019 09:26:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2019	Water	K2674-04	C Flash Point	Cool 4 deg C	PSEG01	K11	SV28545L	05/02/2019	1010A

Date/Time 05/03/19 14:30
 Received by: KS
 Relinquished by: CP

Date/Time 05/03/19 16:05
 Received by: CP
 Relinquished by: KS