


LB102556

245.1

INSTRUMENT ID: CUS

 μ Abs.:

42858



Linear

A= 0.0000e+000

B= 2.3129e-004

C= -9.4722e-003

Rho= 0.9998011

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	30	0.000	30				
0.05	0.050	0.061	0.011	306	0.0 %	306				
0.2	0.200	0.213	0.013	961	0.0 %	961				
2.5	2.500	2.435	-0.065	10569	0.0 %	10569				
5.0	5.000	4.985	-0.015	21594	0.0 %	21594				
7.5	7.500	7.656	0.156	33141	0.0 %	33141				
10.0	10.000	9.903	-0.097	42858	0.0 %	42858				

LB102556 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 May 2019 11:21:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	30	-	0.0000	245.1	PPB	07 May 2019 11:21:50	S	Std
0.05 - 1	50-05	306	-	0.0500	245.1	PPB	07 May 2019 11:24:08	S	Std
0.2 - 1	50-2	961	-	0.2000	245.1	PPB	07 May 2019 11:26:25	S	Std
2.5 - 1	52-5	10569	-	2.5000	245.1	PPB	07 May 2019 11:28:42	S	Std
5.0 - 1	55-0	21594	-	5.0000	245.1	PPB	07 May 2019 11:30:59	S	Std
7.5 - 1	57-5	33141	-	7.5000	245.1	PPB	07 May 2019 11:33:15	S	Std
10.0 - 1	510-0	42858	-	10.0000	245.1	PPB	07 May 2019 11:35:31	S	Std
ICV30 - 1	ICV30	16834	3.8840	-245.1		PPB	07 May 2019 11:45:14	U	SMPL
ICB30 - 1	ICB30	54	0.0030	-245.1		PPB	07 May 2019 11:47:29	U	SMPL
CCV42 - 1	CCV42	21614	4.9895	-245.1		PPB	07 May 2019 11:49:44	U	SMPL
CCB42 - 1	CCB42	-8	-0.0113	-245.1		PPB	07 May 2019 11:52:00	U	SMPL
CRA - 1	CRA	296	0.0590	-245.1		PPB	07 May 2019 11:54:15	U	SMPL
HighStd - 1	HighStd	42729	9.8731	-245.1		PPB	07 May 2019 11:56:30	U	SMPL
ChkStd - 1	ChkStd	30154	6.9647	-245.1		PPB	07 May 2019 11:58:46	U	SMPL
PB119464BL - 1	PBW	-19	-0.0139	-245.1		PPB	07 May 2019 12:01:03	U	SMPL
PB119464BS - 1	LCSW	17822	4.1125	-245.1		PPB	07 May 2019 12:03:19	U	SMPL
K2550-01 - 1	55-ST-10	55	0.0032	-245.1		PPB	07 May 2019 12:05:35	U	SMPL
K2550-01DUP - 1	55-ST-10DUP	97	0.0130	-245.1		PPB	07 May 2019 12:07:52	U	SMPL
K2550-01MS - 1	55-ST-10MS	19436	4.4858	-245.1		PPB	07 May 2019 12:10:09	U	SMPL
K2550-01MSD - 1	55-ST-10MSD	19276	4.4488	-245.1		PPB	07 May 2019 12:12:26	U	SMPL
K2597-02 - 1	OB-722	42	0.0002	-245.1		PPB	07 May 2019 12:14:42	U	SMPL
CCV43 - 1	CCV43	21580	4.9817	-245.1		PPB	07 May 2019 12:16:57	U	SMPL
CCB43 - 1	CCB43	-14	-0.0127	-245.1		PPB	07 May 2019 12:19:13	U	SMPL
K2655-01 - 1	COMP	202	0.0372	-245.1		PPB	07 May 2019 12:21:28	U	SMPL
K2550-01LX5 - 1		-31	-0.0166	-245.1		PPB	07 May 2019 12:23:44	U	SMPL
K2550-01A - 1		19521	4.5055	-245.1		PPB	07 May 2019 12:25:59	U	SMPL
CCV44 - 1	CCV44	21261	4.9079	-245.1		PPB	07 May 2019 12:28:16	U	SMPL
CCB44 - 1	CCB44	71	0.0069	-245.1		PPB	07 May 2019 12:33:43	U	SMPL