



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102566

Slope : 97.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/06/2019	14:10
2	CAL2	1	Water	NA	NA	21.1	7.01	05/06/2019	14:11
3	CAL3	1	Water	NA	NA	21.1	10.03	05/06/2019	14:12
4	ICV	1	Water	NA	NA	21.2	6.99	05/06/2019	14:14
5	CCV1	1	Water	NA	NA	21.1	2.02	05/06/2019	14:15
6	K2698-01	1	Solid	20.02	20	22.5	7.31	05/06/2019	14:16
7	K2698-01DUP	1	Solid	20.03	20	22.6	7.32	05/06/2019	14:19
8	CCV2	1	Water	NA	NA	21.1	12.02	05/06/2019	14:20

LB 102566

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2698

WorkList ID : 125705

Date : 05-06-2019 14:00:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2019	Solid	K2698-01 - 17	pH	Cool 4 deg C	TULL01	K41	STOCKPILE	05/02/2019	9045D

Date/Time 05.06.19 14:05
Received by: 
Relinquished by: 

Date/Time 05.06.19 15:10
Received by: 
Relinquished by: 