



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102574

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/06/2019	15:02
2	K2680-01		1	8	100.00	>212.0	>212.0	05/06/2019	15:32
3	K2680-01DUP		1	8	100.00	>212.0	>212.0	05/06/2019	15:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 102579

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP5619

WorkList ID : 125704

Date : 05-06-2019 13:38:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2019	Solid	K2680-01	B Flash Point	Cool 4 deg C	USAR03	K31	SSSI-0502-S-TCLP-8	05/02/2019	1010A

Date/Time 05/06/19 14:40
Received by: KS
Relinquished by: CP

Date/Time 05/06/19 15:55
Received by: CP
Relinquished by: KS