



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB102575

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	05/06/2019	16:20
2	K2696-02		1	9	100.00	>212.0	>212.0	05/06/2019	16:45
3	K2696-02DUP		1	9	100.00	>212.0	>212.0	05/06/2019	17:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 102575

WorkList Name : FP5619-1

WorkList ID : 125707

Date : 05-06-2019 16:33:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/10/2019	Water	K2696-02	Flash Point	Cool 4 deg C	PARS02	K31	WC-20190502	05/02/2019	1010A

Date/Time 05/06/19 16:00
 Received by: KS
 Relinquished by: CP

Date/Time 05/06/19 17:15
 Received by: CP
 Relinquished by: KS