



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102583

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/07/2019	11:15
2	CAL2	1	Water	NA	NA	20.1	7.00	05/07/2019	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	05/07/2019	11:17
4	ICV	1	Water	NA	NA	20.1	6.99	05/07/2019	11:18
5	CCV1	1	Water	NA	NA	20.1	2.02	05/07/2019	11:20
6	K2713-01	1	Water	NA	NA	19.6	7.91	05/07/2019	11:21
7	K2713-01DUP	1	Water	NA	NA	19.7	7.92	05/07/2019	11:22
8	K2713-02	1	Water	NA	NA	19.5	8.84	05/07/2019	11:24
9	CCV2	1	Water	NA	NA	20.2	12.02	05/07/2019	11:25

LB 102583

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2713

WorkList ID : 125741

Date : 05-07-2019 11:06:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2019	Water	K2713-01 ✓	pH	Cool 4 deg C	PSEG01	K21	NF1133	05/07/2019	9040C
05/07/2019	Water	K2713-02 ✓	pH	Cool 4 deg C	PSEG01	K21	N48962	05/07/2019	9040C

Date/Time 05/07/19 11:10
Received by: Jn
Relinquished by: CW

Date/Time 05/07/19 12:00
Received by: CW
Relinquished by: Jn