



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102600

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/07/2019	18:00
2	CAL2	1	Water	NA	NA	21.1	7.01	05/07/2019	18:02
3	CAL3	1	Water	NA	NA	21.1	10.04	05/07/2019	18:03
4	ICV	1	Water	NA	NA	21.2	6.99	05/07/2019	18:05
5	CCV1	1	Water	NA	NA	21.1	2.02	05/07/2019	18:06
6	K2641-02	1	Solid	20.02	20	24.4	8.30	05/07/2019	18:08
7	K2641-02DUP	1	Solid	20.03	20	24.3	8.31	05/07/2019	18:09
8	K2641-04	1	Solid	20.05	20	24.4	8.68	05/07/2019	18:10
9	K2641-06	1	Solid	20.02	20	24.2	8.32	05/07/2019	18:11
10	K2641-08	1	Solid	20.04	20	24.3	9.05	05/07/2019	18:12
11	K2641-10	1	Solid	20.02	20	24.2	8.92	05/07/2019	18:14
12	K2643-02	1	Solid	20.03	20	24.2	8.65	05/07/2019	18:15
13	K2643-04	1	Solid	20.02	20	24.3	8.52	05/07/2019	18:16
14	K2645-02	1	Solid	20.02	20	25.0	8.18	05/07/2019	18:19
15	K2645-04	1	Solid	20.03	20	24.9	8.15	05/07/2019	18:20
16	CCV2	1	Water	NA	NA	21.2	12.02	05/07/2019	18:21
17	K2645-06	1	Solid	20.05	20	24.9	8.20	05/07/2019	18:22
18	K2645-08	1	Solid	20.03	20	24.7	7.70	05/07/2019	18:24
19	K2645-10	1	Solid	20.01	20	24.7	8.60	05/07/2019	18:25
20	K2703-04	1	Solid	20.03	20	23.0	8.16	05/07/2019	18:26

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	K2728-05	1	Solid	20.02	20	24.0	8.55	05/07/2019	18:27
22	K2728-05DUP	1	Solid	20.01	20	24.1	8.56	05/07/2019	18:29
23	CCV3	1	Water	NA	NA	21.1	2.02	05/07/2019	18:30

LD102600

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K2699

WorkList ID : 125750

Date : 05-07-2019 13:16:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2019	Solid	K2641-02	Corrosivity	Cool 4 deg C	PSEG05	K32	CL-02-050719-A	05/07/2019	9045D
05/07/2019	Solid	K2641-04	Corrosivity	Cool 4 deg C	PSEG05	K32	CL-02-050719-B	05/07/2019	9045D
05/07/2019	Solid	K2641-06	Corrosivity	Cool 4 deg C	PSEG05	K32	CL-02-050719-C	05/07/2019	9045D
05/07/2019	Solid	K2641-08	Corrosivity	Cool 4 deg C	PSEG05	K32	CL-02-050719-D	05/07/2019	9045D
05/07/2019	Solid	K2641-10	Corrosivity	Cool 4 deg C	PSEG05	K32	CL-02-050719-E	05/07/2019	9045D
05/07/2019	Solid	K2643-02	Corrosivity	Cool 4 deg C	PSEG05	K32	OR-03-050719-A	05/07/2019	9045D
05/07/2019	Solid	K2643-04	Corrosivity	Cool 4 deg C	PSEG05	K32	OR-03-050719-B	05/07/2019	9045D
05/07/2019	Solid	K2645-02	Corrosivity	Cool 4 deg C	PSEG05	K33	PL-02-050719-A	05/07/2019	9045D
05/07/2019	Solid	K2645-04	Corrosivity	Cool 4 deg C	PSEG05	K33	PL-02-050719-B	05/07/2019	9045D
05/07/2019	Solid	K2645-06	Corrosivity	Cool 4 deg C	PSEG05	K33	PL-02-050719-C	05/07/2019	9045D
05/07/2019	Solid	K2645-08	Corrosivity	Cool 4 deg C	PSEG05	K33	PL-02-050719-D	05/07/2019	9045D
05/07/2019	Solid	K2645-10	Corrosivity	Cool 4 deg C	PSEG05	K33	PL-02-050719-E	05/07/2019	9045D
05/03/2019	Solid	K2703-04	Corrosivity	Cool 4 deg C	WSP001		P-3-OW(0-10)	05/03/2019	9045D
05/07/2019	Solid	K2728-05	Corrosivity	Cool 4 deg C	WSP001	K33	P-3A-OW(0-10)	05/07/2019	9045D

Date/Time 05.07.19 16:35
 Received by: JD
 Relinquished by: CO

Date/Time 05.07.19 18:30
 Received by: CO
 Relinquished by: JD