

7471B


LB102609

INSTRUMENT ID : CU1

Linear ▼

 μ Abs.:

60622



A= 0.0000e+000

B= 1.6788e-004

C= -3.1201e-002

Rho= 0.9993080

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	73	0.000	73				
0.2	0.200	0.185	-0.015	1288	0.0 %	1288				
2.5	2.500	2.635	0.135	15880	0.0 %	15880				
5.0	5.000	5.022	0.022	30102	0.0 %	30102				
7.5	7.500	7.231	-0.269	43257	0.0 %	43257				
10.0	10.000	10.146	0.146	60622	0.0 %	60622				

LB102609 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 May 2019 11:27:48

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	73	-	0.00007471B	PPB	08 May 2019 11:28:56	S	Std
0.2 - 1	50-2	1288	-	0.20007471B	PPB	08 May 2019 11:31:14	S	Std
2.5 - 1	52-5	15880	-	2.50007471B	PPB	08 May 2019 11:33:32	S	Std
5.0 - 1	55-0	30102	-	5.00007471B	PPB	08 May 2019 11:35:49	S	Std
7.5 - 1	57-5	43257	-	7.50007471B	PPB	08 May 2019 11:38:05	S	Std
10.0 - 1	510-0	60622	-	10.00007471B	PPB	08 May 2019 11:40:22	S	Std
ICV32 - 1	ICV32	24715	4.1180	-7471B	PPB	08 May 2019 11:43:41	U	SMPL
ICB32 - 1	ICB32	-10	-0.0329	-7471B	PPB	08 May 2019 11:45:57	U	SMPL
CCV52 - 1	CCV52	30614	5.1083	-7471B	PPB	08 May 2019 11:48:13	U	SMPL
CCB52 - 1	CCB52	-22	-0.0349	-7471B	PPB	08 May 2019 11:50:28	U	SMPL
CRA - 1	CRA	1284	0.1844	-7471B	PPB	08 May 2019 11:52:44	U	SMPL
HighStd - 1	HighStd	61958	10.3703	-7471B	PPB	08 May 2019 11:54:59	U	SMPL
ChkStd - 1	ChkStd	40324	6.7384	-7471B	PPB	08 May 2019 11:57:15	U	SMPL
PB119541BL - 1	PBS	-115	-0.0505	-7471B	PPB	08 May 2019 11:59:32	U	SMPL
PB119541BS - 1	LCSS	22387	3.7271	-7471B	PPB	08 May 2019 12:01:51	U	SMPL
K2639-01 - 1	NB-08-050319-A	1227	0.1748	-7471B	PPB	08 May 2019 12:04:07	U	SMPL
K2639-01DUP - 1	NB-08-050319-ADUP	1255	0.1795	-7471B	PPB	08 May 2019 12:06:24	U	SMPL
K2639-01MS - 1	NB-08-050319-AMS	25069	4.1774	-7471B	PPB	08 May 2019 12:08:41	U	SMPL
K2639-01MSD - 1	NB-08-050319-AMSD	25075	4.1784	-7471B	PPB	08 May 2019 12:10:58	U	SMPL
K2639-03 - 1	NB-08-050319-B	2513	0.3907	-7471B	PPB	08 May 2019 12:13:15	U	SMPL
CCV53 - 1	CCV53	30557	5.0987	-7471B	PPB	08 May 2019 12:15:31	U	SMPL
CCB53 - 1	CCB53	-31	-0.0364	-7471B	PPB	08 May 2019 12:17:46	U	SMPL
K2641-01 - 1	CL-02-050719-A	2854	0.4479	-7471B	PPB	08 May 2019 12:20:03	U	SMPL
K2641-03 - 1	CL-02-050719-B	4076	0.6531	-7471B	PPB	08 May 2019 12:22:19	U	SMPL
K2641-05 - 1	CL-02-050719-C	3781	0.6036	-7471B	PPB	08 May 2019 12:24:34	U	SMPL
K2641-07 - 1	CL-02-050719-D	2769	0.4337	-7471B	PPB	08 May 2019 12:26:50	U	SMPL
K2641-09 - 1	CL-02-050719-E	18939	3.1483	-7471B	PPB	08 May 2019 12:29:06	U	SMPL
K2643-01 - 1	OR-03-050719-A	5029	0.8131	-7471B	PPB	08 May 2019 12:31:23	U	SMPL
K2643-03 - 1	OR-03-050719-B	11406	1.8836	-7471B	PPB	08 May 2019 12:33:39	U	SMPL
K2698-01 - 1	STOCKPILE	18119	3.0106	-7471B	PPB	08 May 2019 12:35:57	U	SMPL
K2703-02 - 1	P-3-OW(0-10)	56708	9.4890	-7471B	PPB	08 May 2019 12:38:15	U	SMPL
K2704-02 - 1	SB-2(24-27)	3040	0.4792	-7471B	PPB	08 May 2019 12:40:31	U	SMPL
CCV54 - 1	CCV54	30729	5.1276	-7471B	PPB	08 May 2019 12:42:48	U	SMPL
CCB54 - 1	CCB54	-14	-0.0336	-7471B	PPB	08 May 2019 12:45:05	U	SMPL
K2704-04 - 1	SB-2(33.5-35.5)	255	0.0116	-7471B	PPB	08 May 2019 12:47:22	U	SMPL
K2704-06 - 1	SB-1(25-25.5)	714	0.0887	-7471B	PPB	08 May 2019 12:49:38	U	SMPL
K2704-08 - 1	SB-1(35-37)	194	0.0014	-7471B	PPB	08 May 2019 12:51:53	U	SMPL
K2728-02 - 1	P-3A-OW(0-10)	37502	6.2646	-7471B	PPB	08 May 2019 12:54:09	U	SMPL
K2729-01 - 1	B-1	2585	0.4028	-7471B	PPB	08 May 2019 12:56:25	U	SMPL
K2729-02 - 1	B-3	908	0.1212	-7471B	PPB	08 May 2019 12:58:43	U	SMPL
K2729-03 - 1	B-4	33691	5.6249	-7471B	PPB	08 May 2019 13:00:59	U	SMPL
K2729-04 - 1	B-5	922	0.1236	-7471B	PPB	08 May 2019 13:03:16	U	SMPL
PB119542BL - 1	PBS	-22	-0.0349	-7471B	PPB	08 May 2019 13:05:35	U	SMPL
PB119542BS - 1	LCSS	21351	3.5532	-7471B	PPB	08 May 2019 13:07:51	U	SMPL
CCV55 - 1	CCV55	31211	5.2085	-7471B	PPB	08 May 2019 13:10:06	U	SMPL
CCB55 - 1	CCB55	-42	-0.0383	-7471B	PPB	08 May 2019 13:12:23	U	SMPL
K2645-01 - 1	PL-02-050719-A	1606	0.2384	-7471B	PPB	08 May 2019 13:14:39	U	SMPL
K2645-01DUP - 1	PL-02-050719-ADUP	1572	0.2327	-7471B	PPB	08 May 2019 13:16:55	U	SMPL
K2645-01MS - 1	PL-02-050719-AMS	23939	3.9877	-7471B	PPB	08 May 2019 13:19:11	U	SMPL
K2645-01MSD - 1	PL-02-050719-AMSD	23735	3.9534	-7471B	PPB	08 May 2019 13:21:27	U	SMPL
K2645-03 - 1	PL-02-050719-B	1081	0.1503	-7471B	PPB	08 May 2019 13:23:43	U	SMPL
K2645-05 - 1	PL-02-050719-C	1373	0.1993	-7471B	PPB	08 May 2019 13:26:00	U	SMPL
K2645-07 - 1	PL-02-050719-D	2268	0.3496	-7471B	PPB	08 May 2019 13:28:16	U	SMPL
K2645-09 - 1	PL-02-050719-E	1333	0.1926	-7471B	PPB	08 May 2019 13:30:33	U	SMPL
PB119543BL - 1	PB119543BL	-13	-0.0334	-7471B	PPB	08 May 2019 13:32:51	U	SMPL
PB119543BS - 1	PB119543BS	21311	3.5465	-7471B	PPB	08 May 2019 13:35:07	U	SMPL
CCV56 - 1	CCV56	30993	5.1719	-7471B	PPB	08 May 2019 13:37:23	U	SMPL
CCB56 - 1	CCB56	-76	-0.0440	-7471B	PPB	08 May 2019 13:39:39	U	SMPL
K2551-11 - 1	PT-MET-SOIL	345763	58.0156	-7471B	PPB	08 May 2019 13:41:57	U	SMPL
K2551-11 - 1	PT-MET-SOIL	346967	58.2178	-7471B	PPB	08 May 2019 13:44:13	U	SMPL
K2639-01LX5 - 1		-652	-0.1407	-7471B	PPB	08 May 2019 13:46:39	U	SMPL
K2639-01A - 1		23449	3.9054	-7471B	PPB	08 May 2019 13:49:08	U	SMPL
K2645-01LX5 - 1		-7	-0.0324	-7471B	PPB	08 May 2019 13:51:24	U	SMPL
K2645-01A - 1		22480	3.7427	-7471B	PPB	08 May 2019 13:53:41	U	SMPL
K2551-11DLX10 - 1		37661	6.2913	-7471B	PPB	08 May 2019 14:05:41	U	SMPL
K2551-11DLX10 - 1		37775	6.3105	-7471B	PPB	08 May 2019 14:07:58	U	SMPL
CCV57 - 1	CCV57	30440	5.0791	-7471B	PPB	08 May 2019 14:10:16	U	SMPL
CCB57 - 1	CCB57	-124	-0.0520	-7471B	PPB	08 May 2019 14:12:34	U	SMPL