

LB102622


ISM02.4

INSTRUMENT ID : CU1

Linear

 μ Abs. :

40854



A= 0.0000e+000

B= 2.4649e-004

C= -4.1451e-002

Rho= 0.9997983

Accept = Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.014	-0.014	112	0.000	112					1
0.05	0.050										
0.20	0.200	0.178	-0.022	892	0.0 %	892					-11
2.50	2.500	2.632	0.132	10847	0.0 %	10847					5
5.00	5.000	4.882	-0.118	19974	0.0 %	19974					-2
7.50	7.500	7.493	-0.007	30566	0.0 %	30566					0
10.0	10.000	10.029	0.029	40854	0.0 %	40854					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	112	-		0 ISM02.4	PPB	5/8/2019 15:21	Std
	0.2 S0.2	892	-		0.2 ISM02.4	PPB	5/8/2019 15:23	Std
	2.5 S2.5	10847	-		2.5 ISM02.4	PPB	5/8/2019 15:25	Std
	5 S5.0	19974	-		5 ISM02.4	PPB	5/8/2019 15:27	Std
	7.5 S7.5	30566	-		7.5 ISM02.4	PPB	5/8/2019 15:30	Std
	10 S10.0	40854	-		10 ISM02.4	PPB	5/8/2019 15:32	Std
ICV	ICV	16998	4.1484 -		ISM02.4	PPB	5/8/2019 15:35	SMPL
ICB0	ICB	68	-0.0247 -		ISM02.4	PPB	5/8/2019 15:37	SMPL
CCV058	CCV058	20569	5.0286 -		ISM02.4	PPB	5/8/2019 15:39	SMPL
CCB058	CCB058	-15	-0.0451 -		ISM02.4	PPB	5/8/2019 15:42	SMPL
PB119552BL	PBW552	59	-0.0269 -		ISM02.4	PPB	5/8/2019 15:44	SMPL
K2688-22	MA5169	-1	-0.0417 -		ISM02.4	PPB	5/8/2019 15:46	SMPL
PB119553BL	PBW552	50	-0.0291 -		ISM02.4	PPB	5/8/2019 15:48	SMPL
K2473-04	MA5143	-3	-0.0422 -		ISM02.4	PPB	5/8/2019 15:51	SMPL
CCV059	CCV059	19948	4.8755 -		ISM02.4	PPB	5/8/2019 15:53	SMPL
CCB059	CCB059	-1	-0.0417 -		ISM02.4	PPB	5/8/2019 15:55	SMPL