


LB102623

ISM02.4

INSTRUMENT ID : CU1

Linear



A= 0.0000e+000

B= 2.0097e-004

C= -2.5144e-002

Rho= 0.9986592

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.031	-0.031	-31	0.000	-31					—
0.05	0.050										
0.20	0.200	0.187	-0.013	1055	0.0 %	1055					-7
2.50	2.500	2.628	0.128	13204	0.0 %	13204					5
5.00	5.000	5.135	0.135	25678	0.0 %	25678					3
7.50	7.500	7.108	-0.392	35496	0.0 %	35496					-5
10.0	10.000	10.172	0.172	50741	0.0 %	50741					2

ps 05.08.19

LB102623

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-31	-		0 ISM02.4	PPB	5/8/2019 16:41	Std
	0.2 S0.2	1055	-		0.2 ISM02.4	PPB	5/8/2019 16:44	Std
	2.5 S2.5	13204	-		2.5 ISM02.4	PPB	5/8/2019 16:46	Std
	5 S5.0	25678	-		5 ISM02.4	PPB	5/8/2019 16:48	Std
	7.5 S7.5	35496	-		7.5 ISM02.4	PPB	5/8/2019 16:50	Std
	10 S10.0	50741	-		10 ISM02.4	PPB	5/8/2019 16:55	Std
ICV	ICV	21027	4.2006 -		ISM02.4	PPB	5/8/2019 16:59	SMPL
ICB	ICB	38	-0.0175 -		ISM02.4	PPB	5/8/2019 17:01	SMPL
CCV060	CCV060	25390	5.0775 -		ISM02.4	PPB	5/8/2019 17:03	SMPL
CCB060	CCB060	-8	-0.0268 -		ISM02.4	PPB	5/8/2019 17:06	SMPL
PB119554BL	PBS554	25	-0.0201 -		ISM02.4	PPB	5/8/2019 17:08	SMPL
K2688-01	MA5104	40755	8.1654 -		ISM02.4	PPB	5/8/2019 17:10	SMPL
K2688-06	MA5155	-63	-0.0378 -		ISM02.4	PPB	5/8/2019 17:15	SMPL
K2688-07	MA5156	27	-0.0197 -		ISM02.4	PPB	5/8/2019 17:17	SMPL
K2688-10	MA5159	26	-0.0199 -		ISM02.4	PPB	5/8/2019 17:19	SMPL
K2688-13	MA5162	32	-0.0187 -		ISM02.4	PPB	5/8/2019 17:21	SMPL
K2688-16	MA5165	62	-0.0127 -		ISM02.4	PPB	5/8/2019 17:24	SMPL
K2688-18	MA5167	170	0.009 -		ISM02.4	PPB	5/8/2019 17:26	SMPL
K2688-19	MA5167D	147	0.0044 -		ISM02.4	PPB	5/8/2019 17:28	SMPL
K2688-20	MA5167S	14716	2.9323 -		ISM02.4	PPB	5/8/2019 17:35	SMPL
PB119555BL	PBS555	10	-0.0231 -		ISM02.4	PPB	5/8/2019 17:37	SMPL
K2473-01	MA5105	12528	2.4926 -		ISM02.4	PPB	5/8/2019 17:39	SMPL
K2473-02	MA5139	86	-0.0079 -		ISM02.4	PPB	5/8/2019 17:44	SMPL
K2473-03	MA5140	79	-0.0093 -		ISM02.4	PPB	5/8/2019 17:46	SMPL
K2473-05	MA5145	31	-0.0189 -		ISM02.4	PPB	5/8/2019 17:48	SMPL
K2473-06	MA5152	19	-0.0213 -		ISM02.4	PPB	5/8/2019 17:50	SMPL
K2473-07	MA5152D	29	-0.0193 -		ISM02.4	PPB	5/8/2019 17:53	SMPL
K2473-08	MA5152S	13902	2.7687 -		ISM02.4	PPB	5/8/2019 17:55	SMPL
CCV061	CCV061	25408	5.0811 -		ISM02.4	PPB	5/8/2019 17:57	SMPL
CCB061	CCB061	27	-0.0197 -		ISM02.4	PPB	5/8/2019 18:00	SMPL