



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102624

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/08/2019	17:10
2	CAL2	1	Water	NA	NA	21.1	7.00	05/08/2019	17:11
3	CAL3	1	Water	NA	NA	21.1	10.03	05/08/2019	17:12
4	ICV	1	Water	NA	NA	21.2	6.99	05/08/2019	17:14
5	CCV1	1	Water	NA	NA	21.1	2.02	05/08/2019	17:15
6	K2727-01	1	Solid	20.02	20	23.7	6.91	05/08/2019	17:16
7	K2727-02	1	Solid	20.03	20	23.5	6.63	05/08/2019	17:17
8	K2746-02	1	Solid	20.04	20	21.7	9.40	05/08/2019	17:19
9	K2750-01	1	Solid	20.02	20	22.9	7.40	05/08/2019	17:20
10	K2750-05	1	Solid	20.01	20	23.4	8.30	05/08/2019	17:22
11	K2754-06	1	Solid	20.02	20	23.3	7.93	05/08/2019	17:24
12	K2754-08	1	Solid	20.05	20	22.8	8.11	05/08/2019	17:25
13	K2754-08DUP	1	Solid	20.03	20	22.9	8.12	05/08/2019	17:26
14	CCV2	1	Water	NA	NA	21.1	12.01	05/08/2019	17:30

1B102624

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K2750

WorkList ID : 125800

Date : 05-08-2019 16:13:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2019	Solid	K2727-01	Corrosivity	Cool 4 deg C	CLEA11		OWS-2	05/02/2019	9045D
05/02/2019	Solid	K2727-02	Corrosivity	Cool 4 deg C	CLEA11		OWS-3	05/02/2019	9045D
05/08/2019	Solid	K2746-02	Corrosivity	Cool 4 deg C	TECT01	K51	COMP-5.8.19	05/08/2019	9045D
05/08/2019	Solid	K2750-01	Corrosivity	Cool 4 deg C	PSEG01	K42	VB16199	05/08/2019	9045D
05/08/2019	Solid	K2750-05	Corrosivity	Cool 4 deg C	PSEG01	K42	72-11995	05/08/2019	9045D
05/08/2019	Solid	K2754-06	Corrosivity	Cool 4 deg C	PSEG01	K43	C-1	05/08/2019	9045D
05/08/2019	Solid	K2754-08	Corrosivity	Cool 4 deg C	PSEG01	K43	C-2	05/08/2019	9045D

Date/Time 05-08-19 16:20
 Received by: ji
 Relinquished by: ji

Date/Time 05-08-19 17:35
 Received by: ji
 Relinquished by: ji