



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102625

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/08/2019	17:10
2	CAL2	1	Water	NA	NA	21.1	7.00	05/08/2019	17:11
3	CAL3	1	Water	NA	NA	21.1	10.03	05/08/2019	17:12
4	ICV	1	Water	NA	NA	21.2	6.99	05/08/2019	17:14
5	CCV1	1	Water	NA	NA	21.2	2.02	05/08/2019	17:35
6	K2751-06	1	Water	NA	NA	23.0	5.34	05/08/2019	17:36
7	K2751-06DUP	1	Water	NA	NA	23.1	5.35	05/08/2019	17:37
8	K2751-07	1	Water	NA	NA	23.1	5.32	05/08/2019	17:39
9	K2751-08	1	Water	NA	NA	23.0	5.89	05/08/2019	17:40
10	K2751-09	1	Water	NA	NA	22.7	4.71	05/08/2019	17:41
11	K2751-10	1	Water	NA	NA	22.8	5.27	05/08/2019	17:42
12	K2751-11	1	Water	NA	NA	23.0	5.34	05/08/2019	17:43
13	K2751-12	1	Water	NA	NA	23.0	5.32	05/08/2019	17:45
14	K2751-13	1	Water	NA	NA	23.1	6.91	05/08/2019	17:46
15	K2751-14	1	Water	NA	NA	23.0	6.80	05/08/2019	17:48
16	CCV2	1	Water	NA	NA	21.1	12.02	05/08/2019	17:50
17	K2751-15	1	Water	NA	NA	23.3	10.20	05/08/2019	17:51
18	K2751-16	1	Water	NA	NA	22.8	8.36	05/08/2019	17:53
19	K2751-19	1	Water	NA	NA	23.3	6.92	05/08/2019	18:00
20	CCV3	1	Water	NA	NA	21.1	2.02	05/08/2019	18:02

LB 102625

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K2751 W

WorkList ID : 125801

Date : 05-08-2019 16:14:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/08/2019	Water	K2751-06 ✓	pH	Cool 4 deg C	PSEG01	K42	30108	05/08/2019	9040C
05/08/2019	Water	K2751-07 ✓	pH	Cool 4 deg C	PSEG01	K42	30109	05/08/2019	9040C
05/08/2019	Water	K2751-08 ✓	pH	Cool 4 deg C	PSEG01	K42	30117	05/08/2019	9040C
05/08/2019	Water	K2751-09 ✓	pH	Cool 4 deg C	PSEG01	K42	30118	05/08/2019	9040C
05/08/2019	Water	K2751-10 ✓	pH	Cool 4 deg C	PSEG01	K42	30129	05/08/2019	9040C
05/08/2019	Water	K2751-11 ✓	pH	Cool 4 deg C	PSEG01	K42	30163	05/08/2019	9040C
05/08/2019	Water	K2751-12 ✓	pH	Cool 4 deg C	PSEG01	K42	30131	05/08/2019	9040C
05/08/2019	Water	K2751-13 ✓	pH	Cool 4 deg C	PSEG01	K42	30132	05/08/2019	9040C
05/08/2019	Water	K2751-14 ✓	pH	Cool 4 deg C	PSEG01	K42	30133	05/08/2019	9040C
05/08/2019	Water	K2751-15 ✓	pH	Cool 4 deg C	PSEG01	K42	30128	05/08/2019	9040C
05/08/2019	Water	K2751-16 ✓	pH	Cool 4 deg C	PSEG01	K42	30152	05/08/2019	9040C
05/08/2019	Water	K2751-19 ✓	pH	Cool 4 deg C	PSEG01	K42	30121	05/08/2019	9040C

Date/Time 050819 16.25
Received by: Jy
Relinquished by: _____

Date/Time 050819 18.00
Received by:
Relinquished by: