



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: AMANDEEP

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB102628

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP73610
COD calibration std. 100 ppm	WP73607
COD calibration std. 10 ppm	WP73614
COD calibration std. 150 ppm	WP73609
COD calibration std. 50 ppm	WP73608
COD calibration std. 0 ppm	WP73605
COD ICV-LCS std, 50ppm	WP74417
COD CCV std, 50ppm	WP74418
COD Digestion Vials Low Level 0-150Mg/L	W2471

Temp In (C): 150	Date In: 05/08/2019	Time In: 11:49
Temp Out (C): 150	Date Out: 05/08/2019	Time Out: 13:49

Intercept: -0.3388

Slope: 1.0087

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	04/12/2019	14:30
2	CAL2	10	1	Water	10.000	04/12/2019	14:31
3	CAL3	50	1	Water	49.000	04/12/2019	14:32
4	CAL4	100	1	Water	101.000	04/12/2019	14:33
5	CAL5	150	1	Water	151.000	04/12/2019	14:34



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: AMANDEEP

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB102628

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.905	04/12/2019	14:35
2	ICB				1	Water	0.000	0.336	04/12/2019	14:35
3	CCV1	50			1	Water	51.000	50.896	05/08/2019	14:36
4	CCB1				1	Water	0.000	0.336	05/08/2019	14:37
5	LB102628BL				1	Water	0.000	0.336	05/08/2019	14:38
6	LB102628BS	50			1	Water	54.000	53.870	05/08/2019	14:39
7	K2736-02				1	Water	10.000	10.250	05/08/2019	14:40
8	K2736-02DUP				1	Water	11.000	11.241	05/08/2019	14:41
9	K2736-02MS	50			1	Water	58.000	57.836	05/08/2019	14:42
10	K2736-02MSD	50			1	Water	58.000	57.836	05/08/2019	14:43
11	CCV2	50			1	Water	50.000	49.905	05/09/2019	14:45
12	CCB2				1	Water	0.000	0.336	05/09/2019	14:46

LB 102628

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD050819

WorkList ID : 125791

Date : 05-08-2019 11:19:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/21/2019	Water	K2736-02	COD	Conc H2SO4 1%	SHAW22	K31	OUTFALL001-190507	05/07/2019	SM5220 D

Date/Time 5-18/19 11:25
 Received by: AK
 Relinquished by: AK

Date/Time 5/8/19 12:30
 Received by: AK
 Relinquished by: AK