

LB102633


7470A

INSTRUMENT ID : CU1

Linear

 μ Abs. :

44677



A= 0.0000e+000

B= 2.2527e-004

C= 1.5173e-002

Rho= 0.9998595

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	9	0.000	9				
0.2	0.200	0.204	0.004	840	0.0 %	840				
2.5	2.500	2.500	-0.000	11029	0.0 %	11029				
5.0	5.000	5.025	0.025	22239	0.0 %	22239				
7.5	7.500	7.375	-0.125	32670	0.0 %	32670				
10.0	10.000	10.079	0.079	44677	0.0 %	44677				

LB102633 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 May 2019 12:26:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30.0	9	-	0.0000	7470A	PPB	09 May 2019 13:09:48	S	Std
0.2 - 1	50.2	840	-	0.2000	7470A	PPB	09 May 2019 13:11:55	S	Std
2.5 - 1	52.5	11029	-	2.5000	7470A	PPB	09 May 2019 13:14:04	S	Std
5.0 - 1	55.0	22239	-	5.0000	7470A	PPB	09 May 2019 13:16:11	S	Std
7.5 - 1	57.5	32670	-	7.5000	7470A	PPB	09 May 2019 13:18:19	S	Std
10.0 - 1	510.0	44677	-	10.0000	7470A	PPB	09 May 2019 13:20:25	S	Std
ICV35 - 1	ICV35	17102	3.8677	-	7470A	PPB	09 May 2019 13:24:41	U	SMPL
ICB35 - 1	ICB35	-26	0.0093	-	7470A	PPB	09 May 2019 13:26:47	U	SMPL
CCV62 - 1	CCV62	22135	5.0014	-	7470A	PPB	09 May 2019 13:28:53	U	SMPL
CCB62 - 1	CCB62	-45	0.0050	-	7470A	PPB	09 May 2019 13:30:59	U	SMPL
CRA - 1	CRA	868	0.2107	-	7470A	PPB	09 May 2019 13:33:05	U	SMPL
HighStd - 1	HighStd	44071	9.9428	-	7470A	PPB	09 May 2019 13:35:11	U	SMPL
ChkStd - 1	ChkStd	30045	6.7833	-	7470A	PPB	09 May 2019 13:37:16	U	SMPL
PB119575BL - 1	PBW	-88	-0.0047	-	7470A	PPB	09 May 2019 13:39:24	U	SMPL
PB119575BS - 1	LCSW	17987	4.0670	-	7470A	PPB	09 May 2019 13:41:31	U	SMPL
K2641-02 - 1	CL-02-050719-A	-36	0.0071	-	7470A	PPB	09 May 2019 13:43:38	U	SMPL
K2641-02DUP - 1	CL-02-050719-ADUP	1	0.0154	-	7470A	PPB	09 May 2019 13:45:46	U	SMPL
K2641-02MS - 1	CL-02-050719-AMS	17979	4.0652	-	7470A	PPB	09 May 2019 13:47:54	U	SMPL
K2641-02MSD - 1	CL-02-050719-AMSD	18268	4.1303	-	7470A	PPB	09 May 2019 13:50:01	U	SMPL
K2641-04 - 1	CL-02-050719-B	-11	0.0127	-	7470A	PPB	09 May 2019 13:52:08	U	SMPL
CCV63 - 1	CCV63	22335	5.0465	-	7470A	PPB	09 May 2019 13:54:13	U	SMPL
CCB63 - 1	CCB63	-46	0.0048	-	7470A	PPB	09 May 2019 13:56:20	U	SMPL
K2641-06 - 1	CL-02-050719-C	48	0.0260	-	7470A	PPB	09 May 2019 13:58:26	U	SMPL
K2641-08 - 1	CL-02-050719-D	24	0.0206	-	7470A	PPB	09 May 2019 14:00:33	U	SMPL
K2641-10 - 1	CL-02-050719-E	16	0.0188	-	7470A	PPB	09 May 2019 14:02:39	U	SMPL
K2643-02 - 1	OR-03-050719-A	15	0.0186	-	7470A	PPB	09 May 2019 14:04:46	U	SMPL
K2643-04 - 1	OR-03-050719-B	7	0.0167	-	7470A	PPB	09 May 2019 14:06:53	U	SMPL
K2645-02 - 1	PL-02-050719-A	0	0.0152	-	7470A	PPB	09 May 2019 14:09:00	U	SMPL
K2645-04 - 1	PL-02-050719-B	-25	0.0095	-	7470A	PPB	09 May 2019 14:11:07	U	SMPL
K2645-06 - 1	PL-02-050719-C	4	0.0161	-	7470A	PPB	09 May 2019 14:13:15	U	SMPL
K2645-08 - 1	PL-02-050719-D	5	0.0163	-	7470A	PPB	09 May 2019 14:15:23	U	SMPL
K2645-10 - 1	PL-02-050719-E	19	0.0195	-	7470A	PPB	09 May 2019 14:17:29	U	SMPL
CCV64 - 1	CCV64	21831	4.9329	-	7470A	PPB	09 May 2019 14:19:36	U	SMPL
CCB64 - 1	CCB64	-9	0.0131	-	7470A	PPB	09 May 2019 14:21:42	U	SMPL
K2698-02 - 1	STOCKPILE	39	0.0240	-	7470A	PPB	09 May 2019 14:23:48	U	SMPL
K2703-04 - 1	P-3-OW(0-10)	31	0.0222	-	7470A	PPB	09 May 2019 14:25:56	U	SMPL
K2727-01 - 1	OWS-2	35	0.0231	-	7470A	PPB	09 May 2019 14:28:02	U	SMPL
K2727-02 - 1	OWS-3	31	0.0222	-	7470A	PPB	09 May 2019 14:30:09	U	SMPL
K2728-05 - 1	P-3A-OW(0-10)	23	0.0204	-	7470A	PPB	09 May 2019 14:32:15	U	SMPL
PB119577BL - 1	PBW	-4	0.0143	-	7470A	PPB	09 May 2019 14:34:23	U	SMPL
PB119577BS - 1	LCSW	18445	4.1702	-	7470A	PPB	09 May 2019 14:36:30	U	SMPL
K2696-01 - 1	WC-20190502	190	0.0580	-	7470A	PPB	09 May 2019 14:38:38	U	SMPL
K2696-01DUP - 1	WC-20190502DUP	191	0.0582	-	7470A	PPB	09 May 2019 14:40:46	U	SMPL
K2696-01MS - 1	WC-20190502MS	19171	4.3337	-	7470A	PPB	09 May 2019 14:42:53	U	SMPL
CCV65 - 1	CCV65	22791	5.1492	-	7470A	PPB	09 May 2019 14:44:59	U	SMPL
CCB65 - 1	CCB65	-78	-0.0024	-	7470A	PPB	09 May 2019 14:47:06	U	SMPL
K2696-01MSD - 1	WC-20190502MSD	19199	4.3400	-	7470A	PPB	09 May 2019 14:49:13	U	SMPL
PB119585BL - 1	PBW	-60	0.0017	-	7470A	PPB	09 May 2019 14:51:20	U	SMPL
PB119585BS - 1	LCSW	18038	4.0785	-	7470A	PPB	09 May 2019 14:53:27	U	SMPL
K2746-02 - 1	COMP-5.8.19	286	0.0796	-	7470A	PPB	09 May 2019 14:55:33	U	SMPL
K2746-02DUP - 1	COMP-5.8.19DUP	342	0.0922	-	7470A	PPB	09 May 2019 14:57:40	U	SMPL
K2746-02MS - 1	COMP-5.8.19MS	17406	3.9361	-	7470A	PPB	09 May 2019 14:59:47	U	SMPL
K2746-02MSD - 1	COMP-5.8.19MSD	17723	4.0075	-	7470A	PPB	09 May 2019 15:01:54	U	SMPL
K2750-05 - 1	72-11995	415	0.1087	-	7470A	PPB	09 May 2019 15:04:03	U	SMPL
K2751-02 - 1	COMP-3	670	0.1661	-	7470A	PPB	09 May 2019 15:06:11	U	SMPL
K2754-06 - 1	C-1	649	0.1614	-	7470A	PPB	09 May 2019 15:08:18	U	SMPL
CCV66 - 1	CCV66	22932	5.1810	-	7470A	PPB	09 May 2019 15:10:24	U	SMPL
CCB66 - 1	CCB66	-44	0.0053	-	7470A	PPB	09 May 2019 15:12:31	U	SMPL
K2754-08 - 1	C-2	517	0.1316	-	7470A	PPB	09 May 2019 15:14:38	U	SMPL
PB119586BL - 1	PBW	16	0.0188	-	7470A	PPB	09 May 2019 15:16:45	U	SMPL
PB119586BS - 1	LCSW	18288	4.1348	-	7470A	PPB	09 May 2019 15:18:51	U	SMPL
K2751-04 - 1	COMP-1	276	0.0773	-	7470A	PPB	09 May 2019 15:20:57	U	SMPL
K2751-04DUP - 1	COMP-1DUP	296	0.0819	-	7470A	PPB	09 May 2019 15:23:04	U	SMPL
K2751-04MS - 1	COMP-1MS	19823	4.4806	-	7470A	PPB	09 May 2019 15:25:11	U	SMPL
K2751-04MSD - 1	COMP-1MSD	19815	4.4788	-	7470A	PPB	09 May 2019 15:27:19	U	SMPL
K2751-18 - 1	COMP-2	601	0.1506	-	7470A	PPB	09 May 2019 15:29:26	U	SMPL
K2752-02 - 1	COMP-4	548	0.1386	-	7470A	PPB	09 May 2019 15:31:35	U	SMPL
PB119575TB - 1	PB119520TB	37	0.0235	-	7470A	PPB	09 May 2019 15:33:42	U	SMPL
CCV67 - 1	CCV67	23201	5.2416	-	7470A	PPB	09 May 2019 15:35:49	U	SMPL
CCB67 - 1	CCB67	-55	0.0028	-	7470A	PPB	09 May 2019 15:37:57	U	SMPL
PB119585TB - 1	PB119560TB	46	0.0255	-	7470A	PPB	09 May 2019 15:47:14	U	SMPL
PB119586TB - 1	PB119562TB	132	0.0449	-	7470A	PPB	09 May 2019 15:49:22	U	SMPL
K2641-02LX5 - 1		-34	0.0075	-	7470A	PPB	09 May 2019 15:55:49	U	SMPL
K2641-02A - 1		18170	4.1082	-	7470A	PPB	09 May 2019 15:57:56	U	SMPL
K2696-01LX5 - 1		-33	0.0077	-	7470A	PPB	09 May 2019 16:00:03	U	SMPL
K2696-01A - 1		19049	4.3063	-	7470A	PPB	09 May 2019 16:02:10	U	SMPL
K2746-02LX5 - 1		-5	0.0140	-	7470A	PPB	09 May 2019 16:04:17	U	SMPL
K2746-02A - 1		17355	3.9247	-	7470A	PPB	09 May 2019 16:06:25	U	SMPL
K2751-04LX5 - 1		31	0.0222	-	7470A	PPB	09 May 2019 16:08:33	U	SMPL
K2751-04A - 1		20153	4.5549	-	7470A	PPB	09 May 2019 16:10:41	U	SMPL
CCV68 - 1	CCV68	21794	4.9246	-	7470A	PPB	09 May 2019 16:12:48	U	SMPL

LB102633 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 May 2019 12:26:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB68 - 1	CCB68	-32	0.0080		7470A	PPB	09 May 2019 16:14:55	U	SMPL