



## Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102639

| Reagent/Standard                        | Lot/Log # |
|-----------------------------------------|-----------|
| Residual chlorine std, 0.8PPM           | WP74555   |
| Residual chlorine std, 0PPM             | WP74551   |
| Residual chlorine std, 0.1PPM           | WP74552   |
| Residual chlorine std, 0.2PPM           | WP74553   |
| Residual chlorine std, and CCV , 0.4PPM | WP74554   |
| Residual chlorine std, 1.6PPM           | WP74556   |
| Residual chlorine ICV-LCS, 0.4PPM       | WP74557   |
| Total Chlorine Powder Pillows           | W2249     |

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

| Seq | Lab ID | True Val<br>(mg/l) | Dilution | Initial<br>Reading | Final<br>Reading | Difference | AnalDate   | Anal Time |
|-----|--------|--------------------|----------|--------------------|------------------|------------|------------|-----------|
| 1   | CAL1   | 0                  | 1        | 0.000              | 0.000            | 0.000      | 05/09/2019 | 12:45     |
| 2   | CAL2   | 0.1                | 1        | 0.000              | 0.100            | 0.100      | 05/09/2019 | 12:45     |
| 3   | CAL3   | 0.2                | 1        | 0.000              | 0.200            | 0.200      | 05/09/2019 | 12:46     |
| 4   | CAL4   | 0.4                | 1        | 0.000              | 0.400            | 0.400      | 05/09/2019 | 12:47     |
| 5   | CAL5   | 0.8                | 1        | 0.000              | 0.800            | 0.800      | 05/09/2019 | 12:48     |
| 6   | CAL6   | 1.6                | 1        | 0.000              | 1.600            | 1.600      | 05/09/2019 | 12:49     |



## Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB102639

| Seq | Lab ID      | Initial Weight | Final Vol | True Value (mg/L) | DF | Initial Reading | Final Reading | Diff.  | Result (mg/L) | Anal Date  | Anal Time |
|-----|-------------|----------------|-----------|-------------------|----|-----------------|---------------|--------|---------------|------------|-----------|
| 1   | ICV         |                |           | 0.4               | 1  | 0.0000          | 0.3900        | 0.3900 | 0.3900        | 05/09/2019 | 12:50     |
| 2   | ICB         |                |           |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 05/09/2019 | 12:50     |
| 3   | CCV1        |                |           | 0.4               | 1  | 0.0000          | 0.4000        | 0.4000 | 0.4000        | 05/09/2019 | 12:51     |
| 4   | CCB1        |                |           |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 05/09/2019 | 12:52     |
| 5   | LB102639BL  | 50             | 50        |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 05/09/2019 | 12:54     |
| 6   | LB102639BS  | 50             | 50        | 0.4               | 1  | 0.0000          | 0.4000        | 0.4000 | 0.4000        | 05/09/2019 | 12:55     |
| 7   | K2781-01    | 50             | 50        |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 05/09/2019 | 12:56     |
| 8   | K2781-01DUP | 50             | 50        |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 05/09/2019 | 12:57     |
| 9   | K2781-01MS  | 50             | 50        | 0.4               | 1  | 0.0000          | 0.3500        | 0.3500 | 0.3500        | 05/09/2019 | 12:58     |
| 10  | K2781-01MSD | 50             | 50        | 0.4               | 1  | 0.0000          | 0.3600        | 0.3600 | 0.3600        | 05/09/2019 | 12:59     |
| 11  | CCV2        |                |           | 0.4               | 1  | 0.0000          | 0.4000        | 0.4000 | 0.4000        | 05/09/2019 | 13:00     |
| 12  | CCB2        |                |           |                   | 1  | 0.0000          | 0.0000        | 0.0000 | 0.0000        | 05/09/2019 | 13:01     |

CB 102039

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCL050919

WorkList ID : 125837

Date : 05-09-2019 12:17:22

| Due Date   | Matrix | Sample   | Test              | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method      |
|------------|--------|----------|-------------------|--------------|----------|------------------|-----------------|--------------|-------------|
| 05/08/2019 | Water  | K2781-01 | Residual Chlorine | Cool 4 deg C | SHAW22   | O11              | OUTFALL001-WK-2 | 05/08/2019   | SM4500 Cl G |

Date/Time 05-09-19 12:28  
 Received by: AD  
 Relinquished by: G

Date/Time 05-09-19 13:15  
 Received by: G  
 Relinquished by: AD