



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB102642

Ambient Barometric Pressure (mmHg): 755.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	05/09/2019	09:45
2	K2751-03		1	8	100.00	>212.0	>212.0	05/09/2019	10:20
3	K2751-03DUP		1	8	100.00	>212.0	>212.0	05/09/2019	10:50
4	K2751-19		1	9	100.00	>212.0	>212.0	05/09/2019	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB102642

WorkList Name : FP5919

WorkList ID : 125825

Date : 05-09-2019 09:15:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/15/2019	Water	K2751-03 C	Flash Point	Cool 4 deg C	PSEG01	K42	COMP-1	05/08/2019	1010A
05/15/2019	Water	K2751-19 C	Flash Point	Cool 4 deg C	PSEG01	K42	30121	05/08/2019	1010A

Date/Time 05/09/19 09:20
 Received by: KS
 Relinquished by: CP

Date/Time 05/09/19 11:25
 Received by: CP
 Relinquished by: KS