



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB102645

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/09/2019	13:10
2	CAL2	1	Water	NA	NA	20.1	7.00	05/09/2019	13:11
3	CAL3	1	Water	NA	NA	20.1	10.02	05/09/2019	13:12
4	ICV	1	Water	NA	NA	20.2	6.99	05/09/2019	13:15
5	CCV1	1	Water	NA	NA	20.1	2.02	05/09/2019	13:16
6	K2771-01	1	Solid	20.02	20	23.7	5.20	05/09/2019	13:17
7	K2774-01	1	Solid	20.04	20	22.9	5.63	05/09/2019	13:19
8	CCV2	1	Water	NA	NA	20.2	12.02	05/09/2019	13:20

LB 102645

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-K2771 WorkList ID : 125840 Date : 05-09-2019 12:54:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/09/2019	Solid	K2771-01 -17	pH	Cool 4 deg C	TULL02	K31	FK-PS-01-047	05/09/2019	9045D
05/09/2019	Solid	K2774-01 -17	pH	Cool 4 deg C	TULL02	O11	FK-PS-01-048	05/09/2019	9045D

Date/Time 05-09-19 10:35
Received by: Jo
Relinquished by: C8

Date/Time 05-09-19 14:25
Received by: C8
Relinquished by: Jo