



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/10/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 05/09/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 05/10/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB102653

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K2692-01	A5170	1	1.00	2.00	2.00	100.0
K2692-02	A5171	2	1.00	2.00	2.00	100.0
K2692-03	A5172	3	1.14	9.55	9.09	94.5
K2692-04	A5173	4	1.16	9.85	8.86	88.6
K2692-05	A5174	5	1.12	9.59	8.81	90.8
K2692-06	VHBLK01	6	1.00	2.00	2.00	100.0
K2692-07	A5107	7	1.00	2.00	2.00	100.0
K2692-08	A5113	8	1.00	2.00	2.00	100.0
K2692-09	A5118	9	1.00	2.00	2.00	100.0
K2692-10	A5175	10	1.14	9.93	8.69	85.9
K2692-11	A5175MS	11	1.14	9.93	8.69	85.9
K2692-12	A5175MSD	12	1.14	9.93	8.69	85.9
K2692-13	A5176	13	1.14	9.6	8.5	87.0
K2692-14	A5177	14	1.18	9.54	9.02	93.8
K2692-15	A5178	15	1.15	9.94	9.24	92.0
K2692-16	A5179	16	1.17	9.97	9.18	91.0
K2692-17	A5180	17	1.1	9.65	8.95	91.8
K2692-18	A5181	18	1.13	9.79	9.33	94.7
K2692-19	A5182	19	1.11	9.85	9.31	93.8
K2692-20	A5183	20	1.18	9.98	9.37	93.1
K2692-21	A5184	21	1.17	9.7	9.22	94.4
K2692-22	A5185	22	1.15	9.73	8.68	87.8
K2692-23	A5186	23	1.14	9.96	8.77	86.5
K2692-24	A5186MS	24	1.14	9.96	8.77	86.5
K2692-25	A5186MSD	25	1.14	9.96	8.77	86.5
K2692-26	A5187	26	1.11	9.95	8.93	88.5
K2692-27	A5189	27	1.15	9.95	9.45	94.3
K2692-28	A5193	28	1.19	9.88	8.99	89.8

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$