



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB102662
Analysis Date: 05/10/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 05/10/2019
Extraction IN Time: 07:40
Extraction OUT Time: 08:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB102662BL	LB102662BL	WATER	1.3	1000	100	2.1033	2.1033	3.02	2.1033	2.1033	0.0000	0
2	LB102662BS	LB102662BS	WATER	1.3	1000	100	2.8574	2.8574	3.01	2.8764	2.8764	0.0190	19
3	LB102662BSD	LB102662BSD	WATER	1.3	1000	100	2.1639	2.1639	3.03	2.1831	2.1831	0.0192	19.2
4	K2789-01	SEWER-OUTLET-01	WATER	1.3	1000	100	3.1192	3.1192	3.05	3.1253	3.1253	0.0061	6.1
5	K2789-02	SEWER-OUTLET-02	WATER	1.3	1000	100	2.9572	2.9572	3.04	2.9622	2.9622	0.0050	5
6	K2800-01	EFFL-4	WATER	1.3	1000	100	3.1318	3.1318	3.03	3.1329	3.1329	0.0011	1.1



QC Batch# LB102662 Test: Non-Polar Mate Analysis Date: 05/10/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2515
pH Paper 0-14	M4240
Sodium Sulfate	EP1944
1:1 HCL	WP74116
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	10.00 ML	WP74118
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:41</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:40</u>		
Bal Check Time:	<u>08:00</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:15</u>
		Out Time:	<u>09:40</u>		

After Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:16</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>10:40</u>		
Bal Check Time:	<u>12:25</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:50</u>
		Out Time:	<u>11:15</u>		

29 102662

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPAR K2789

WorkList ID : 125860

Date : 05-09-2019 18:10:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2019	Water	K2789-01-65	Non-Polar Material	Conc H2SO4 tr	TRIS02	O13	SEWER-OUTLET-01	05/09/2019	1664A
05/23/2019	Water	K2789-02-65	Non-Polar Material	Conc H2SO4 tr	TRIS02	O13	SEWER-OUTLET-02	05/09/2019	1664A
05/14/2019	Water	K2800-01-65	Non-Polar Material	Conc H2SO4 tr	AREC01	O11	EFFL-4	05/09/2019	1664A

Date/Time 05.10.19 7:30
 Received by: JB
 Relinquished by: CJ

Date/Time 05.10.19 10:10
 Received by: CJ
 Relinquished by: JB