



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB102677

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/09/2019	17:20
2	CAL2	1	Water	NA	NA	20.1	7.00	05/09/2019	17:21
3	CAL3	1	Water	NA	NA	20.1	10.03	05/09/2019	17:22
4	ICV	1	Water	NA	NA	20.1	6.99	05/09/2019	17:24
5	CCV1	1	Water	NA	NA	20.2	2.02	05/09/2019	17:25
6	K2795-02	1	Water	NA	NA	18.9	7.92	05/09/2019	17:26
7	K2795-02DUP	1	Water	NA	NA	18.8	7.93	05/09/2019	17:27
8	CCV2	1	Water	NA	NA	20.2	12.02	05/09/2019	17:31

20102677

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K2795

WorkList ID : 125923

Date : 05-10-2019 14:43:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/09/2019	Water	K2795-02- ¹⁸	Corrosivity	Cool 4 deg C	PSEG01	K31	TRUCK-803	05/09/2019	9040C

Date/Time 05-09-19 17:15
 Received by: SO
 Relinquished by: SO

Date/Time 05-09-19 18:00
 Received by: SO
 Relinquished by: SO